

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110118\
 Data File : P0050934.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2018 8:03
 Operator : SM/SJ
 Sample : J5759-21
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-6

Manual Integrations
 APPROVED

Sohil
 11/2/2018 6:24:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 09:27:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 30 07:48:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.353	3.348	5827019	3349617	11.657	15.411 #
2)	SA Decachlor...	10.009	8.092	3837571	2246297	12.346m	8.782 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110118\
Data File : PO050934.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2018 8:03
Operator : SM/SJ
Sample : J5759-21
Misc :
ALS Vial : 68 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-6


Manual Integrations
APPROVED

Sohil
11/2/2018 6:24:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 09:27:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 30 07:48:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.353 min
Delta R.T.: -0.003 min
Response: 5827019
Conc: 11.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6


Manual Integrations
APPROVED

Sohil
11/2/2018 6:24:01 PM




#1 Tetrachloro-m-xylene

R.T.: 3.348 min
Delta R.T.: 0.000 min
Response: 3349617
Conc: 15.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.009 min
Delta R.T.: -0.008 min
Response: 3837571
Conc: 12.35 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.092 min
Delta R.T.: -0.004 min
Response: 2246297
Conc: 8.78 ng/ml