

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110122\
 Data File : PO090059.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2022 10:45
 Operator : YP/AJ
 Sample : N5336-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-01-16.5-17.0

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 04:19:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.305	3.483	78608881	25998354	24.459	24.462
2) SA Decachlor...	10.052	8.494	58765903	23726532	23.981	23.037

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110122\
Data File : PO090059.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2022 10:45
Operator : YP/AJ
Sample : N5336-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-01-16.5-17.0

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 04:19:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.305 min
Delta R.T.: -0.002 min
Response: 78608881
Conc: 24.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-16.5-17.0




#1 Tetrachloro-m-xylene

R.T.: 3.483 min
Delta R.T.: -0.004 min
Response: 25998354
Conc: 24.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: -0.001 min
Response: 58765903
Conc: 23.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.494 min
Delta R.T.: -0.002 min
Response: 23726532
Conc: 23.04 ng/ml