

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110123\
 Data File : PO099265.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2023 19:02
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 00:51:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.469	3.625	36441263	14215761	18.443	19.753
2)	SA Decachlor...	10.265	8.619	23716359	10534342	20.935	19.634
Target Compounds							
3)	L1 AR-1016-1	5.645	4.714	49559	144438	0.900	6.450 #
4)	L1 AR-1016-2	5.680	4.745	67320	22201	0.814	0.710
5)	L1 AR-1016-3	5.748	4.949f	12241	71091	0.238	4.198 #
6)	L1 AR-1016-4	5.830	4.949	26646	71091	0.656	4.917 #
7)	L1 AR-1016-5	6.145	5.166	72571	48193	1.710	2.580 #
8)	L2 AR-1221-1	4.680	3.855	135313	114969	5.450	12.376 #
9)	L2 AR-1221-2	4.778	3.924	989672	321502	53.498	49.586
10)	L2 AR-1221-3	4.857	4.018	116721	32696	2.154	1.616
11)	L3 AR-1232-1	4.857	4.018	116721	32696	2.551	1.907 #
12)	L3 AR-1232-2	5.379	4.745	23926	22201	1.013	1.449 #
13)	L3 AR-1232-3	5.680	4.949f	67320	71091	1.705	8.885 #
14)	L3 AR-1232-4	5.830	4.997	26646	119088	1.397	15.479 #
15)	L3 AR-1232-5	5.921	5.166	245092	48193	14.713	5.666 #
16)	L4 AR-1242-1	5.645	4.714	49559	144438	1.145	8.221 #
17)	L4 AR-1242-2	5.680	4.745	67320	22201	1.051	0.916
18)	L4 AR-1242-3	5.748	4.949f	12241	71091	0.304	5.545 #
19)	L4 AR-1242-4	5.830	4.997	26646	119088	0.852	8.460 #
20)	L4 AR-1242-5	6.575	5.534	172381	748124	5.588	44.768 #
21)	L5 AR-1248-1	5.645	4.714	49559	144438	1.459	10.421 #
22)	L5 AR-1248-2	5.921	4.949	245092	71091	4.683	3.579
23)	L5 AR-1248-3	6.145	4.997	72571	119088	1.274	5.711 #
24)	L5 AR-1248-4	6.536	5.166	143893	48193	2.639	1.960 #
25)	L5 AR-1248-5	6.575	5.574	172381	389218	3.035	17.921 #
26)	L6 AR-1254-1	6.502	5.534	369462	748124	5.613	21.114 #
27)	L6 AR-1254-2	6.732	5.658	35467	238338	0.368	7.452 #
28)	L6 AR-1254-3	7.093	6.057	568519	364150	6.360	7.449
29)	L6 AR-1254-4	7.374	6.256f	25419	119661	0.464	4.566 #
30)	L6 AR-1254-5	7.802	6.718	61707	82659	0.916	1.961 #
31)	L7 AR-1260-1	7.261	6.191	51945	299256	0.725	8.483 #
32)	L7 AR-1260-2	7.521	6.360	87065	297730	1.104	7.427 #
33)	L7 AR-1260-3	7.891	0.000	70228	0	1.286	N.D. #
34)	L7 AR-1260-4	8.105	7.059f	53496	97286	0.854	3.367 #
35)	L7 AR-1260-5	8.431	7.273	46008	39969	0.430	0.664 #
36)	L8 AR-1262-1	7.863	6.741	24831	73315	0.274	1.563 #
37)	L8 AR-1262-2	8.431	7.059f	46008	97286	0.345	2.343 #
38)	L8 AR-1262-3	8.731	7.539	30440	107535	0.310	3.366 #
39)	L8 AR-1262-4	8.828	7.539f	67208	107535	0.839	1.956 #
40)	L8 AR-1262-5	9.503	8.056	28715	19212	0.653	0.765
41)	L9 AR-1268-1	8.731	7.539	30440	107535	0.177	1.206 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110123\
 Data File : PO099265.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2023 19:02
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 00:51:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.850	7.539f	85175	107535	0.543	1.362 #
43) L9	AR-1268-3	9.070	0.000	57778	0	0.412	N.D. #
44) L9	AR-1268-4	9.503	8.056f	28715	19212	0.573	0.697
45) L9	AR-1268-5	9.927	8.370	74178	64575	0.181	0.339 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110123\
Data File : PO099265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2023 19:02
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 00:51:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.469 min
Delta R.T.: -0.009 min
Response: 36441263
Conc: 18.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: -0.007 min
Response: 14215761
Conc: 19.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.265 min
Delta R.T.: -0.015 min
Response: 23716359
Conc: 20.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.619 min
Delta R.T.: -0.012 min
Response: 10534342
Conc: 19.63 ng/ml



#3 AR-1016-1


R.T.: 5.645 min
Delta R.T.: -0.006 min
Response: 49559
Conc: 0.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.714 min
Delta R.T.: 0.000 min
Response: 144438
Conc: 6.45 ng/ml




#4 AR-1016-2

R.T.: 5.680 min
Delta R.T.: 0.006 min
Response: 67320
Conc: 0.81 ng/ml



#4 AR-1016-2


R.T.: 4.745 min
Delta R.T.: 0.014 min
Response: 22201
Conc: 0.71 ng/ml



#5 AR-1016-3


R.T.: 5.748 min
Delta R.T.: 0.012 min
Response: 12241
Conc: 0.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.949 min
Delta R.T.: 0.041 min
Response: 71091
Conc: 4.20 ng/ml




#6 AR-1016-4

R.T.: 5.830 min
Delta R.T.: -0.005 min
Response: 26646
Conc: 0.66 ng/ml



#6 AR-1016-4


R.T.: 4.949 min
Delta R.T.: 0.000 min
Response: 71091
Conc: 4.92 ng/ml



#7 AR-1016-5


R.T.: 6.145 min
Delta R.T.: 0.014 min
Response: 72571
Conc: 1.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.166 min
Delta R.T.: 0.003 min
Response: 48193
Conc: 2.58 ng/ml




#8 AR-1221-1

R.T.: 4.680 min
Delta R.T.: -0.006 min
Response: 135313
Conc: 5.45 ng/ml



#8 AR-1221-1


R.T.: 3.855 min
Delta R.T.: 0.011 min
Response: 114969
Conc: 12.38 ng/ml



#9 AR-1221-2


R.T.: 4.778 min
Delta R.T.: 0.006 min
Response: 989672
Conc: 53.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.924 min
Delta R.T.: -0.006 min
Response: 321502
Conc: 49.59 ng/ml




#10 AR-1221-3

R.T.: 4.857 min
Delta R.T.: 0.007 min
Response: 116721
Conc: 2.15 ng/ml



#10 AR-1221-3


R.T.: 4.018 min
Delta R.T.: 0.013 min
Response: 32696
Conc: 1.62 ng/ml



#11 AR-1232-1


R.T.: 4.857 min
Delta R.T.: 0.008 min
Response: 116721
Conc: 2.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.018 min
Delta R.T.: 0.013 min
Response: 32696
Conc: 1.91 ng/ml




#12 AR-1232-2

R.T.: 5.379 min
Delta R.T.: -0.002 min
Response: 23926
Conc: 1.01 ng/ml



#12 AR-1232-2


R.T.: 4.745 min
Delta R.T.: 0.015 min
Response: 22201
Conc: 1.45 ng/ml



#13 AR-1232-3


R.T.: 5.680 min
Delta R.T.: 0.008 min
Response: 67320
Conc: 1.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.949 min
Delta R.T.: 0.043 min
Response: 71091
Conc: 8.88 ng/ml




#14 AR-1232-4

R.T.: 5.830 min
Delta R.T.: -0.003 min
Response: 26646
Conc: 1.40 ng/ml



#14 AR-1232-4


R.T.: 4.997 min
Delta R.T.: 0.007 min
Response: 119088
Conc: 15.48 ng/ml



#15 AR-1232-5


R.T.: 5.921 min
Delta R.T.: -0.003 min
Response: 245092
Conc: 14.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.166 min
Delta R.T.: 0.004 min
Response: 48193
Conc: 5.67 ng/ml




#16 AR-1242-1

R.T.: 5.645 min
Delta R.T.: -0.006 min
Response: 49559
Conc: 1.14 ng/ml



#16 AR-1242-1


R.T.: 4.714 min
Delta R.T.: 0.002 min
Response: 144438
Conc: 8.22 ng/ml



#17 AR-1242-2


R.T.: 5.680 min
Delta R.T.: 0.007 min
Response: 67320
Conc: 1.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.745 min
Delta R.T.: 0.015 min
Response: 22201
Conc: 0.92 ng/ml




#18 AR-1242-3

R.T.: 5.748 min
Delta R.T.: 0.013 min
Response: 12241
Conc: 0.30 ng/ml



#18 AR-1242-3


R.T.: 4.949 min
Delta R.T.: 0.043 min
Response: 71091
Conc: 5.55 ng/ml



#19 AR-1242-4


R.T.: 5.830 min
Delta R.T.: -0.004 min
Response: 26646
Conc: 0.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.997 min
Delta R.T.: 0.008 min
Response: 119088
Conc: 8.46 ng/ml




#20 AR-1242-5

R.T.: 6.575 min
Delta R.T.: 0.002 min
Response: 172381
Conc: 5.59 ng/ml



#20 AR-1242-5


R.T.: 5.534 min
Delta R.T.: 0.021 min
Response: 748124
Conc: 44.77 ng/ml



#21 AR-1248-1


R.T.: 5.645 min
Delta R.T.: -0.005 min
Response: 49559
Conc: 1.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.714 min
Delta R.T.: 0.002 min
Response: 144438
Conc: 10.42 ng/ml




#22 AR-1248-2

R.T.: 5.921 min
Delta R.T.: -0.004 min
Response: 245092
Conc: 4.68 ng/ml



#22 AR-1248-2


R.T.: 4.949 min
Delta R.T.: 0.000 min
Response: 71091
Conc: 3.58 ng/ml



#23 AR-1248-3


R.T.: 6.145 min
Delta R.T.: 0.015 min
Response: 72571
Conc: 1.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.997 min
Delta R.T.: 0.007 min
Response: 119088
Conc: 5.71 ng/ml




#24 AR-1248-4

R.T.: 6.536 min
Delta R.T.: 0.001 min
Response: 143893
Conc: 2.64 ng/ml



#24 AR-1248-4


R.T.: 5.166 min
Delta R.T.: 0.004 min
Response: 48193
Conc: 1.96 ng/ml



#25 AR-1248-5


R.T.: 6.575 min
Delta R.T.: 0.002 min
Response: 172381
Conc: 3.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.574 min
Delta R.T.: 0.020 min
Response: 389218
Conc: 17.92 ng/ml




#26 AR-1254-1

R.T.: 6.502 min
Delta R.T.: -0.006 min
Response: 369462
Conc: 5.61 ng/ml



#26 AR-1254-1


R.T.: 5.534 min
Delta R.T.: 0.022 min
Response: 748124
Conc: 21.11 ng/ml



#27 AR-1254-2


R.T.: 6.732 min
Delta R.T.: 0.004 min
Response: 35467
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.658 min
Delta R.T.: -0.001 min
Response: 238338
Conc: 7.45 ng/ml




#28 AR-1254-3

R.T.: 7.093 min
Delta R.T.: -0.002 min
Response: 568519
Conc: 6.36 ng/ml



#28 AR-1254-3


R.T.: 6.057 min
Delta R.T.: -0.006 min
Response: 364150
Conc: 7.45 ng/ml



#29 AR-1254-4


R.T.: 7.374 min
Delta R.T.: -0.006 min
Response: 25419
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.256 min
Delta R.T.: -0.035 min
Response: 119661
Conc: 4.57 ng/ml




#30 AR-1254-5

R.T.: 7.802 min
Delta R.T.: 0.001 min
Response: 61707
Conc: 0.92 ng/ml



#30 AR-1254-5


R.T.: 6.718 min
Delta R.T.: 0.009 min
Response: 82659
Conc: 1.96 ng/ml



#31 AR-1260-1


R.T.: 7.261 min
Delta R.T.: 0.000 min
Response: 51945
Conc: 0.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.191 min
Delta R.T.: -0.004 min
Response: 299256
Conc: 8.48 ng/ml




#32 AR-1260-2

R.T.: 7.521 min
Delta R.T.: 0.004 min
Response: 87065
Conc: 1.10 ng/ml



#32 AR-1260-2


R.T.: 6.360 min
Delta R.T.: -0.023 min
Response: 297730
Conc: 7.43 ng/ml



#33 AR-1260-3


R.T.: 7.891 min
Delta R.T.: 0.013 min
Response: 70228
Conc: 1.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.537 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.105 min
Delta R.T.: 0.000 min
Response: 53496
Conc: 0.85 ng/ml



#34 AR-1260-4


R.T.: 7.059 min
Delta R.T.: 0.051 min
Response: 97286
Conc: 3.37 ng/ml



#35 AR-1260-5


R.T.: 8.431 min
Delta R.T.: 0.002 min
Response: 46008
Conc: 0.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.273 min
Delta R.T.: 0.022 min
Response: 39969
Conc: 0.66 ng/ml




#36 AR-1262-1

R.T.: 7.863 min
Delta R.T.: -0.013 min
Response: 24831
Conc: 0.27 ng/ml



#36 AR-1262-1


R.T.: 6.741 min
Delta R.T.: -0.005 min
Response: 73315
Conc: 1.56 ng/ml



#37 AR-1262-2


R.T.: 8.431 min
Delta R.T.: 0.003 min
Response: 46008
Conc: 0.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.059 min
Delta R.T.: 0.053 min
Response: 97286
Conc: 2.34 ng/ml




#38 AR-1262-3

R.T.: 8.731 min
Delta R.T.: -0.014 min
Response: 30440
Conc: 0.31 ng/ml



#38 AR-1262-3


R.T.: 7.539 min
Delta R.T.: 0.008 min
Response: 107535
Conc: 3.37 ng/ml



#39 AR-1262-4


R.T.: 8.828 min
Delta R.T.: -0.008 min
Response: 67208
Conc: 0.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.539 min
Delta R.T.: -0.056 min
Response: 107535
Conc: 1.96 ng/ml




#40 AR-1262-5

R.T.: 9.503 min
Delta R.T.: -0.003 min
Response: 28715
Conc: 0.65 ng/ml



#40 AR-1262-5


R.T.: 8.056 min
Delta R.T.: -0.035 min
Response: 19212
Conc: 0.76 ng/ml



#41 AR-1268-1


R.T.: 8.731 min
Delta R.T.: -0.014 min
Response: 30440
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.539 min
Delta R.T.: 0.007 min
Response: 107535
Conc: 1.21 ng/ml




#42 AR-1268-2

R.T.: 8.850 min
Delta R.T.: 0.009 min
Response: 85175
Conc: 0.54 ng/ml



#42 AR-1268-2


R.T.: 7.539 min
Delta R.T.: -0.057 min
Response: 107535
Conc: 1.36 ng/ml



#43 AR-1268-3


R.T.: 9.070 min
Delta R.T.: -0.002 min
Response: 57778
Conc: 0.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.802 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.503 min
Delta R.T.: -0.004 min
Response: 28715
Conc: 0.57 ng/ml



#44 AR-1268-4


R.T.: 8.056 min
Delta R.T.: -0.035 min
Response: 19212
Conc: 0.70 ng/ml



#45 AR-1268-5

R.T.: 9.927 min
Delta R.T.: -0.006 min
Response: 74178
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.370 min
Delta R.T.: -0.008 min
Response: 64575
Conc: 0.34 ng/ml