

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110218\
 Data File : P0051001.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Nov 2018 2:16
 Operator : SM/SJ
 Sample : J5776-05
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 NJ-WHITE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 03:49:37 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 30 07:48:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.352	3.347	11197737	4679561	22.402	21.530
2)	SA Decachlor...	10.010	8.090	5359560	4420035	17.243m	17.280

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110218\
Data File : PO051001.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Nov 2018 2:16
Operator : SM/SJ
Sample : J5776-05
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
NJ-WHITE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 03:49:37 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 30 07:48:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.352 min
Delta R.T.: -0.004 min
Response: 11197737
Conc: 22.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
NJ-WHITE




#1 Tetrachloro-m-xylene

R.T.: 3.347 min
Delta R.T.: -0.002 min
Response: 4679561
Conc: 21.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.010 min
Delta R.T.: -0.008 min
Response: 5359560
Conc: 17.24 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.090 min
Delta R.T.: -0.005 min
Response: 4420035
Conc: 17.28 ng/ml