

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110220\
 Data File : PO072971.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2020 12:24
 Operator : DD\AJ
 Sample : L4559-13
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 200FB-005-000.66

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 03:33:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 31 07:06:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.950	4.026	1160555	813680	25.171	23.217
2)	SA Decachlor...	10.985	9.310	657704	523323	17.802	15.692

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110220\
Data File : PO072971.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2020 12:24
Operator : DD\AJ
Sample : L4559-13
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
200FB-005-000.66

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 03:33:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 31 07:06:26 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.950 min
Delta R.T.: 0.000 min
Response: 1160555
Conc: 25.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
200FB-005-000.66




#1 Tetrachloro-m-xylene

R.T.: 4.026 min
Delta R.T.: 0.005 min
Response: 813680
Conc: 23.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.985 min
Delta R.T.: -0.004 min
Response: 657704
Conc: 17.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.310 min
Delta R.T.: -0.006 min
Response: 523323
Conc: 15.69 ng/ml