

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110220\
 Data File : PO072974.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2020 13:15
 Operator : DD\AJ
 Sample : L4557-18
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 200FB-008-0002

Manual Integrations
 APPROVED

Ankita
 11/3/2020 3:33:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 03:35:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 31 07:06:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.951	4.027	1038801	740323	22.530	21.124
2) SA Decachlor...	10.988	9.310	896791	970643	24.274m	29.106m
Target Compounds						
41) L9 AR-1268-1	9.473	8.226	2733032	2345984	371.174	393.431
42) L9 AR-1268-2	9.569	8.290	2946791	2443681	446.916	449.948
43) L9 AR-1268-3	9.796	8.499	754623	608825	135.462	124.898
44) L9 AR-1268-4	10.236	8.783	1811442	2542980	776.763	1164.644 #
45) L9 AR-1268-5	10.651	9.065	3574593	3056449	210.816	212.233

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110220\
Data File : PO072974.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2020 13:15
Operator : DD\AJ
Sample : L4557-18
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
200FB-008-0002


Manual Integrations
APPROVED

Ankita
11/3/2020 3:33:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 03:35:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO102620.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 31 07:06:26 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.951 min
Delta R.T.: 0.000 min
Response: 1038801
Conc: 22.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
200FB-008-0002


Manual Integrations
APPROVED

Ankita
11/3/2020 3:33:32 PM




#1 Tetrachloro-m-xylene

R.T.: 4.027 min
Delta R.T.: 0.005 min
Response: 740323
Conc: 21.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.988 min
Delta R.T.: -0.002 min
Response: 896791
Conc: 24.27 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.310 min
Delta R.T.: -0.006 min
Response: 970643
Conc: 29.11 ng/ml m




#41 AR-1268-1

R.T.: 9.473 min
Delta R.T.: 0.003 min
Response: 2733032
Conc: 371.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
200FB-008-0002


Manual Integrations
APPROVED

Ankita
11/3/2020 3:33:32 PM




#41 AR-1268-1

R.T.: 8.226 min
Delta R.T.: 0.003 min
Response: 2345984
Conc: 393.43 ng/ml




#42 AR-1268-2

R.T.: 9.569 min
Delta R.T.: 0.002 min
Response: 2946791
Conc: 446.92 ng/ml



#42 AR-1268-2

R.T.: 8.290 min
Delta R.T.: 0.002 min
Response: 2443681
Conc: 449.95 ng/ml




#43 AR-1268-3

R.T.: 9.796 min
Delta R.T.: 0.004 min
Response: 754623
Conc: 135.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
200FB-008-0002


Manual Integrations
APPROVED

Ankita
11/3/2020 3:33:32 PM




#43 AR-1268-3

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 608825
Conc: 124.90 ng/ml




#44 AR-1268-4

R.T.: 10.236 min
Delta R.T.: 0.004 min
Response: 1811442
Conc: 776.76 ng/ml



#44 AR-1268-4

R.T.: 8.783 min
Delta R.T.: 0.003 min
Response: 2542980
Conc: 1164.64 ng/ml




#45 AR-1268-5

R.T.: 10.651 min
Delta R.T.: 0.004 min
Response: 3574593
Conc: 210.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
200FB-008-0002

Manual Integrations
APPROVED

Ankita
11/3/2020 3:33:32 PM



#45 AR-1268-5

R.T.: 9.065 min
Delta R.T.: 0.004 min
Response: 3056449
Conc: 212.23 ng/ml