

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0110321\
 Data File : P0082457.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Nov 2021 3:59
 Operator : AJ\MA
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 05:24:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0110321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 03 05:20:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	87.529	86.840 E3	0.8	98	0.00
2 SA Decachlorobiphenyl	63.152	109.137 E3	-72.8#	170#	0.00
41 L9 AR-1268-1	8.163	8.143 E3	0.2	99	0.00
42 L9 AR-1268-2	7.561	7.593 E3	-0.4	98	0.00
43 L9 AR-1268-3	6.561	6.532 E3	0.4	99	0.00
44 L9 AR-1268-4	2.813	2.804 E3	0.3	99	0.00
45 L9 AR-1268-5	22.124	22.208 E3	-0.4	100	0.00

Signal #2

1 SA Tetrachloro-m-xylene	30.947	31.339 E3	-1.3	101	0.00
2 SA Decachlorobiphenyl	28.792	49.796 E3	-73.0#	173#	0.00
41 L9 AR-1268-1	5.230	5.113 E3	2.2	99	0.00
42 L9 AR-1268-2	4.538	4.467 E3	1.6	99	0.00
43 L9 AR-1268-3	3.887	3.771 E3	3.0	98	0.00
44 L9 AR-1268-4	1.589	1.569 E3	1.3	98	0.00
45 L9 AR-1268-5	10.935	10.854 E3	0.7	100	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	2.901	0.000 E3	100.0#	0#	-6.30#
4 L1 AR-1016-2	4.127	0.000 E3	100.0#	0#	-6.33#
5 L1 AR-1016-3	2.561	0.000 E3	100.0#	0#	-6.39#
6 L1 AR-1016-4	2.064	0.000 E3	100.0#	0#	-6.50#
7 L1 AR-1016-5	2.039	0.000 E3	100.0#	0#	-6.82#
8 L2 AR-1221-1	943.061	0.000	100.0#	0#	-5.22#
9 L2 AR-1221-2	648.189	0.000	100.0#	0#	-5.32#
10 L2 AR-1221-3	2.063	0.000 E3	100.0#	0#	-5.41#
11 L3 AR-1232-1	1.732	0.000 E3	100.0#	0#	-5.41#
12 L3 AR-1232-2	884.278	0.000	100.0#	0#	-6.01#
13 L3 AR-1232-3	1.579	0.000 E3	100.0#	0#	-6.33#
14 L3 AR-1232-4	767.930	0.000	100.0#	0#	-6.50#
15 L3 AR-1232-5	552.775	0.000	100.0#	0#	-6.60#
16 L4 AR-1242-1	2.079	0.000 E3	100.0#	0#	-6.30#
17 L4 AR-1242-2	2.917	0.000 E3	100.0#	0#	-6.33#
18 L4 AR-1242-3	1.845	0.000 E3	100.0#	0#	-6.39#
19 L4 AR-1242-4	1.495	0.000 E3	100.0#	0#	-6.50#
20 L4 AR-1242-5	1.496	0.000 E3	100.0#	0#	-7.29#
21 L5 AR-1248-1	1.526	0.000 E3	100.0#	0#	-6.30#
22 L5 AR-1248-2	2.103	0.000 E3	100.0#	0#	-6.60#
23 L5 AR-1248-3	2.426	0.000 E3	100.0#	0#	-6.82#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0110321\
 Data File : P0082457.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Nov 2021 3:59
 Operator : AJ\MA
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 05:24:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0110321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 03 05:20:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	2.569	0.000 E3	100.0#	0#	-7.25#
25 L5 AR-1248-5	2.522	0.000 E3	100.0#	0#	-7.29#
26 L6 AR-1254-1	2.724	0.000 E3	100.0#	0#	-7.22#
27 L6 AR-1254-2	4.166	0.000 E3	100.0#	0#	-7.45#
28 L6 AR-1254-3	4.277	0.000 E3	100.0#	0#	-7.84#
29 L6 AR-1254-4	3.067	0.000 E3	100.0#	0#	-8.13#
30 L6 AR-1254-5	3.216	0.000 E3	100.0#	0#	-8.57#
31 L7 AR-1260-1	3.175	0.000 E3	100.0#	0#	-8.00#
32 L7 AR-1260-2	3.759	0.000 E3	100.0#	0#	-8.27#
33 L7 AR-1260-3	2.832	0.000 E3	100.0#	0#	-8.64#
34 L7 AR-1260-4	3.333	0.000 E3	100.0#	0#	-8.87#
35 L7 AR-1260-5	6.275	0.000 E3	100.0#	0#	-9.21#
36 L8 AR-1262-1	3.710	0.000 E3	100.0#	0#	-8.64#
37 L8 AR-1262-2	6.387	0.000 E3	100.0#	0#	-9.21#
38 L8 AR-1262-3	2.868	0.000 E3	100.0#	0#	-9.52#
39 L8 AR-1262-4	1.688	0.000 E3	100.0#	0#	-9.62#
40 L8 AR-1262-5	2.429	0.000 E3	100.0#	0#	-10.30#

Signal #2

3 L1 AR-1016-1	1.342	0.000 E3	100.0#	0#	-5.22#
4 L1 AR-1016-2	1.841	0.000 E3	100.0#	0#	-5.25#
5 L1 AR-1016-3	996.900	0.000	100.0#	0#	-5.44#
6 L1 AR-1016-4	815.683	0.000	100.0#	0#	-5.49#
7 L1 AR-1016-5	1.019	0.000 E3	100.0#	0#	-5.72#
8 L2 AR-1221-1	381.942	0.000	100.0#	0#	-4.22#
9 L2 AR-1221-2	290.308	0.000	100.0#	0#	-4.32#
10 L2 AR-1221-3	925.322	0.000	100.0#	0#	-4.41#
11 L3 AR-1232-1	789.431	0.000	100.0#	0#	-4.41#
12 L3 AR-1232-2	750.632	0.000	100.0#	0#	-5.24#
13 L3 AR-1232-3	396.553	0.000	100.0#	0#	-5.44#
14 L3 AR-1232-4	345.340	0.000	100.0#	0#	-5.53#
15 L3 AR-1232-5	374.906	0.000	100.0#	0#	-5.72#
16 L4 AR-1242-1	975.889	0.000	100.0#	0#	-5.23#
17 L4 AR-1242-2	1.346	0.000 E3	100.0#	0#	-5.25#
18 L4 AR-1242-3	728.936	0.000	100.0#	0#	-5.44#
19 L4 AR-1242-4	713.912	0.000	100.0#	0#	-5.53#
20 L4 AR-1242-5	923.194	0.000	100.0#	0#	-6.10#
21 L5 AR-1248-1	737.471	0.000	100.0#	0#	-5.22#
22 L5 AR-1248-2	1.018	0.000 E3	100.0#	0#	-5.49#
23 L5 AR-1248-3	1.041	0.000 E3	100.0#	0#	-5.53#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0110321\
 Data File : P0082457.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Nov 2021 3:59
 Operator : AJ\MA
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 05:24:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0110321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 03 05:20:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound			AvgRF	CCRF	%Dev	Area%	Dev(Min)

24	L5	AR-1248-4	1.226	0.000 E3	100.0#	0#	-5.72#
25	L5	AR-1248-5	1.236	0.000 E3	100.0#	0#	-6.14#
26	L6	AR-1254-1	1.942	0.000 E3	100.0#	0#	-6.10#
27	L6	AR-1254-2	1.711	0.000 E3	100.0#	0#	-6.26#
28	L6	AR-1254-3	2.571	0.000 E3	100.0#	0#	-6.68#
29	L6	AR-1254-4	1.818	0.000 E3	100.0#	0#	-6.92#
30	L6	AR-1254-5	2.324	0.000 E3	100.0#	0#	-7.35#
31	L7	AR-1260-1	1.814	0.000 E3	100.0#	0#	-6.82#
32	L7	AR-1260-2	2.393	0.000 E3	100.0#	0#	-7.02#
33	L7	AR-1260-3	2.127	0.000 E3	100.0#	0#	-7.17#
34	L7	AR-1260-4	1.668	0.000 E3	100.0#	0#	-7.66#
35	L7	AR-1260-5	4.164	0.000 E3	100.0#	0#	-7.91#
36	L8	AR-1262-1	1.195	0.000 E3	100.0#	0#	-7.35#
37	L8	AR-1262-2	4.081	0.000 E3	100.0#	0#	-7.91#
38	L8	AR-1262-3	1.631	0.000 E3	100.0#	0#	-8.19#
39	L8	AR-1262-4	2.912	0.000 E3	100.0#	0#	-8.26#
40	L8	AR-1262-5	1.370	0.000 E3	100.0#	0#	-8.76#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0