

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110323\
 Data File : P0099330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2023 09:56
 Operator : YP/AJ
 Sample : PB156850BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB156850BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/06/2023
 Supervised By :mohammad ahmed 11/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 23:00:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.471	3.622	39898082	15469143	20.192	21.495m
2) SA Decachlor...	10.265	8.616	29749599	11973448	26.260m	22.316m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110323\
Data File : PO099330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2023 09:56
Operator : YP/AJ
Sample : PB156850BL
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB156850BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/06/2023
Supervised By :mohammad ahmed 11/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 23:00:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.471 min
Delta R.T.: -0.007 min
Response: 39898082
Conc: 20.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB156850BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/06/2023
Supervised By :mohammad ahmed 11/06/2023




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.010 min
Response: 15469143
Conc: 21.49 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.265 min
Delta R.T.: -0.016 min
Response: 29749599
Conc: 26.26 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.616 min
Delta R.T.: -0.014 min
Response: 11973448
Conc: 22.32 ng/ml m