

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110325\
 Data File : PO114837.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2025 09:44
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/04/2025
 Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 12:51:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 02:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.349	3.651	306.1E6	200.5E6	56.385	50.093
2)	SA Decachlor...	9.930	8.642	202.8E6	182.5E6	55.704	54.757m
Target Compounds							
26)	L6 AR-1254-1	6.330	5.527	99716554	122.6E6	571.594	461.283
27)	L6 AR-1254-2	6.546	5.674	145.8E6	109.0E6	572.517	457.370
28)	L6 AR-1254-3	6.906	6.076	156.1E6	172.7E6	540.360	487.386
29)	L6 AR-1254-4	7.187	6.303	131.8E6	125.3E6	556.834	485.530m
30)	L6 AR-1254-5	7.603	6.721	128.1E6	165.7E6	583.704	517.331

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110325\
Data File : PO114837.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 09:44
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 12:51:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 02:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.349 min
Delta R.T.: -0.006 min
Response: 306110754
Conc: 56.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#1 Tetrachloro-m-xylene

R.T.: 3.651 min
Delta R.T.: -0.009 min
Response: 200486249
Conc: 50.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: -0.005 min
Response: 202823419
Conc: 55.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: -0.029 min
Response: 182516467
Conc: 54.76 ng/ml m




#26 AR-1254-1

R.T.: 6.330 min
Delta R.T.: -0.004 min
Response: 99716554
Conc: 571.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#26 AR-1254-1

R.T.: 5.527 min
Delta R.T.: -0.015 min
Response: 122560274
Conc: 461.28 ng/ml




#27 AR-1254-2

R.T.: 6.546 min
Delta R.T.: -0.004 min
Response: 145823915
Conc: 572.52 ng/ml



#27 AR-1254-2

R.T.: 5.674 min
Delta R.T.: -0.015 min
Response: 108960267
Conc: 457.37 ng/ml




#28 AR-1254-3

R.T.: 6.906 min
Delta R.T.: -0.004 min
Response: 156132025
Conc: 540.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#28 AR-1254-3

R.T.: 6.076 min
Delta R.T.: -0.016 min
Response: 172733033
Conc: 487.39 ng/ml




#29 AR-1254-4

R.T.: 7.187 min
Delta R.T.: -0.003 min
Response: 131772704
Conc: 556.83 ng/ml



#29 AR-1254-4

R.T.: 6.303 min
Delta R.T.: -0.016 min
Response: 125342679
Conc: 485.53 ng/ml m




#30 AR-1254-5

R.T.: 7.603 min
Delta R.T.: -0.003 min
Response: 128108874
Conc: 583.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025



#30 AR-1254-5

R.T.: 6.721 min
Delta R.T.: -0.017 min
Response: 165651320
Conc: 517.33 ng/ml