

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110422\
 Data File : PO090211.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2022 01:29
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 07:37:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.304	3.483	192.3E6	67391242	59.835	63.409
2)	SA Decachlor...	10.049	8.488	144.2E6	60950207	58.859	59.179
Target Compounds							
26)	L6 AR-1254-1	6.341	5.374	54455340	34355365	469.859	583.230
27)	L6 AR-1254-2	6.560	5.523	93110247	31069264	535.321	579.864
28)	L6 AR-1254-3	6.927	5.927	97478989	46207178	561.498	557.501
29)	L6 AR-1254-4	7.215	6.157	70295697	27383485	551.483	551.370
30)	L6 AR-1254-5	7.636	6.577	71629650	42917995	508.518	573.192



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110422\
Data File : PO090211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2022 01:29
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 07:37:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.304 min
Delta R.T.: -0.003 min
Response: 192302948
Conc: 59.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.483 min
Delta R.T.: -0.003 min
Response: 67391242
Conc: 63.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.049 min
Delta R.T.: -0.005 min
Response: 144238576
Conc: 58.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.488 min
Delta R.T.: -0.008 min
Response: 60950207
Conc: 59.18 ng/ml



#26 AR-1254-1


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 54455340
Conc: 469.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.374 min
Delta R.T.: -0.002 min
Response: 34355365
Conc: 583.23 ng/ml




#27 AR-1254-2

R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 93110247
Conc: 535.32 ng/ml



#27 AR-1254-2


R.T.: 5.523 min
Delta R.T.: -0.002 min
Response: 31069264
Conc: 579.86 ng/ml



#28 AR-1254-3


R.T.: 6.927 min
Delta R.T.: 0.000 min
Response: 97478989
Conc: 561.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.927 min
Delta R.T.: -0.003 min
Response: 46207178
Conc: 557.50 ng/ml




#29 AR-1254-4

R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 70295697
Conc: 551.48 ng/ml



#29 AR-1254-4


R.T.: 6.157 min
Delta R.T.: -0.003 min
Response: 27383485
Conc: 551.37 ng/ml



#30 AR-1254-5

R.T.: 7.636 min
Delta R.T.: 0.000 min
Response: 71629650
Conc: 508.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.577 min
Delta R.T.: -0.002 min
Response: 42917995
Conc: 573.19 ng/ml