

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110521\
 Data File : PO082577.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Nov 2021 00:15
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:08:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 03 05:20:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.966	3.961	1846677	667547	21.098	21.571
2)	SA Decachlor...	11.069	9.276	1422091	671256	22.519	23.314
Target Compounds							
8)	L2 AR-1221-1	5.206f	0.000	50698	0	53.759	N.D. #
9)	L2 AR-1221-2	5.322	4.314	22737	8376	35.078	28.853
10)	L2 AR-1221-3	0.000	4.406	0	23342	N.D.	25.225 #
11)	L3 AR-1232-1	0.000	4.406	0	23342	N.D.	29.568 #
14)	L3 AR-1232-4	0.000	5.557f	0	17120	N.D.	49.573 #
19)	L4 AR-1242-4	0.000	5.557f	0	17120	N.D.	23.980 #
20)	L4 AR-1242-5	7.288	6.117f	42741	77886	28.571	84.366 #
23)	L5 AR-1248-3	0.000	5.557f	0	17120	N.D.	16.452 #
24)	L5 AR-1248-4	7.248	0.000	35619	0	13.868	N.D. #
25)	L5 AR-1248-5	7.288	6.117f	42741	77886	16.946	63.031 #
26)	L6 AR-1254-1	7.221	6.117f	449965	77886	165.159	40.099 #
28)	L6 AR-1254-3	7.849	6.665	146644	458	34.287	0.178 #
32)	L7 AR-1260-2	8.298f	0.000	161988	0	43.094	N.D. #
33)	L7 AR-1260-3	0.000	7.167	0	18941	N.D.	8.907 #
38)	L8 AR-1262-3	0.000	8.168f	0	10968	N.D.	6.726 #
41)	L9 AR-1268-1	0.000	8.168f	0	10968	N.D.	2.097 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110521\
Data File : PO082577.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Nov 2021 00:15
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:08:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110321.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 03 05:20:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.966 min
Delta R.T.: -0.004 min
Response: 1846677
Conc: 21.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.961 min
Delta R.T.: -0.005 min
Response: 667547
Conc: 21.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.069 min
Delta R.T.: -0.010 min
Response: 1422091
Conc: 22.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.276 min
Delta R.T.: -0.007 min
Response: 671256
Conc: 23.31 ng/ml



#8 AR-1221-1


R.T.: 5.206 min
Delta R.T.: -0.018 min
Response: 50698
Conc: 53.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.224 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.322 min
Delta R.T.: -0.003 min
Response: 22737
Conc: 35.08 ng/ml



#9 AR-1221-2


R.T.: 4.314 min
Delta R.T.: -0.009 min
Response: 8376
Conc: 28.85 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 5.412 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 4.406 min
Delta R.T.: -0.006 min
Response: 23342
Conc: 25.23 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 5.411 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.406 min
Delta R.T.: -0.006 min
Response: 23342
Conc: 29.57 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 6.502 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#14 AR-1232-4

R.T.: 5.557 min
Delta R.T.: 0.025 min
Response: 17120
Conc: 49.57 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T.: 6.502 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.557 min
Delta R.T.: 0.023 min
Response: 17120
Conc: 23.98 ng/ml



#20 AR-1242-5


R.T.: 7.288 min
Delta R.T.: -0.003 min
Response: 42741
Conc: 28.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 6.117 min
Delta R.T.: 0.018 min
Response: 77886
Conc: 84.37 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 6.819 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.557 min
Delta R.T.: 0.024 min
Response: 17120
Conc: 16.45 ng/ml



#24 AR-1248-4


R.T.: 7.248 min
Delta R.T.: -0.004 min
Response: 35619
Conc: 13.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.717 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 7.288 min
Delta R.T.: -0.003 min
Response: 42741
Conc: 16.95 ng/ml



#25 AR-1248-5


R.T.: 6.117 min
Delta R.T.: -0.024 min
Response: 77886
Conc: 63.03 ng/ml



#26 AR-1254-1


R.T.: 7.221 min
Delta R.T.: 0.000 min
Response: 449965
Conc: 165.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 6.117 min
Delta R.T.: 0.019 min
Response: 77886
Conc: 40.10 ng/ml




#28 AR-1254-3

R.T.: 7.849 min
Delta R.T.: 0.012 min
Response: 146644
Conc: 34.29 ng/ml



#28 AR-1254-3


R.T.: 6.665 min
Delta R.T.: -0.014 min
Response: 458
Conc: 0.18 ng/ml



#32 AR-1260-2


R.T.: 8.298 min
Delta R.T.: 0.027 min
Response: 161988
Conc: 43.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.017 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.640 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.167 min
Delta R.T.: -0.005 min
Response: 18941
Conc: 8.91 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 9.522 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#38 AR-1262-3

R.T.: 8.168 min
Delta R.T.: -0.025 min
Response: 10968
Conc: 6.73 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.524 min
Response: 0
Conc: N.D.



#41 AR-1268-1

R.T.: 8.168 min
Delta R.T.: -0.024 min
Response: 10968
Conc: 2.10 ng/ml