

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110521\
 Data File : PO082546.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Nov 2021 13:54
 Operator : AJ\MA
 Sample : M4475-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB02A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:00:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 03 05:20:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.967	3.963	3652448	1318291	41.728	42.599
2) SA Decachlor...	11.070	9.277	2475396	1135779	39.198	39.447
Target Compounds						
31) L7 AR-1260-1	7.998	6.814	129553	120476	40.798	66.414 #
32) L7 AR-1260-2	8.266	7.012	304575	140057	81.027	58.524 #
33) L7 AR-1260-3	8.634	7.167	151712	132243	53.569	62.185
34) L7 AR-1260-4	8.866	7.651	176997	89347	53.098m	53.567
35) L7 AR-1260-5	9.201	7.900	303228	232799	48.325	55.908



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110521\
Data File : PO082546.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Nov 2021 13:54
Operator : AJ\MA
Sample : M4475-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB02A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:00:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110321.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 03 05:20:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.967 min
Delta R.T.: -0.003 min
Response: 3652448
Conc: 41.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB02A




#1 Tetrachloro-m-xylene

R.T.: 3.963 min
Delta R.T.: -0.003 min
Response: 1318291
Conc: 42.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.070 min
Delta R.T.: -0.009 min
Response: 2475396
Conc: 39.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.277 min
Delta R.T.: -0.006 min
Response: 1135779
Conc: 39.45 ng/ml



#31 AR-1260-1


R.T.: 7.998 min
Delta R.T.: -0.006 min
Response: 129553
Conc: 40.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB02A




#31 AR-1260-1

R.T.: 6.814 min
Delta R.T.: -0.005 min
Response: 120476
Conc: 66.41 ng/ml




#32 AR-1260-2

R.T.: 8.266 min
Delta R.T.: -0.005 min
Response: 304575
Conc: 81.03 ng/ml



#32 AR-1260-2


R.T.: 7.012 min
Delta R.T.: -0.005 min
Response: 140057
Conc: 58.52 ng/ml



#33 AR-1260-3


R.T.: 8.634 min
Delta R.T.: -0.007 min
Response: 151712
Conc: 53.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB02A




#33 AR-1260-3

R.T.: 7.167 min
Delta R.T.: -0.006 min
Response: 132243
Conc: 62.19 ng/ml




#34 AR-1260-4

R.T.: 8.866 min
Delta R.T.: -0.007 min
Response: 176997
Conc: 53.10 ng/ml m



#34 AR-1260-4


R.T.: 7.651 min
Delta R.T.: -0.005 min
Response: 89347
Conc: 53.57 ng/ml



#35 AR-1260-5

R.T.: 9.201 min
Delta R.T.: -0.005 min
Response: 303228
Conc: 48.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB02A



#35 AR-1260-5

R.T.: 7.900 min
Delta R.T.: -0.005 min
Response: 232799
Conc: 55.91 ng/ml