

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110719\
 Data File : P0063547.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Nov 2019 16:18
 Operator : SM/AJ
 Sample : P0110719ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVP0110719

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 04:56:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0110719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 04:54:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	33.090	34.436 E3	-4.1	105	0.00
2 SA Decachlorobiphenyl	54.872	52.970 E3	3.5	99	0.00
3 L1 AR-1016-1	1.387	1.374 E3	0.9	101	0.00
4 L1 AR-1016-2	2.019	2.038 E3	-0.9	103	0.00
5 L1 AR-1016-3	1.269	1.279 E3	-0.8	102	0.00
6 L1 AR-1016-4	1.036	1.050 E3	-1.4	102	0.00
7 L1 AR-1016-5	1.119	1.119 E3	0.0	101	0.00
31 L7 AR-1260-1	2.403	2.346 E3	2.4	100	0.00
32 L7 AR-1260-2	3.020	2.908 E3	3.7	99	0.00
33 L7 AR-1260-3	2.314	2.299 E3	0.6	99	0.00
34 L7 AR-1260-4	2.775	2.711 E3	2.3	99	0.00
35 L7 AR-1260-5	5.548	5.424 E3	2.2	99	0.00

Signal #2

1 SA Tetrachloro-m-xylene	21.085	22.105 E3	-4.8	105	0.00
2 SA Decachlorobiphenyl	29.641	28.840 E3	2.7	99	0.00
3 L1 AR-1016-1	842.305	868.422	-3.1	103	0.00
4 L1 AR-1016-2	1.140	1.180 E3	-3.5	102	0.00
5 L1 AR-1016-3	632.415	637.600	-0.8	102	0.00
6 L1 AR-1016-4	531.232	527.948	0.6	102	0.00
7 L1 AR-1016-5	678.144	702.103	-3.5	103	0.00
31 L7 AR-1260-1	1.414	1.410 E3	0.3	101	0.00
32 L7 AR-1260-2	1.769	1.752 E3	1.0	101	0.00
33 L7 AR-1260-3	1.625	1.620 E3	0.3	101	0.00
34 L7 AR-1260-4	1.411	1.410 E3	0.1	101	0.00
35 L7 AR-1260-5	3.369	3.349 E3	0.6	100	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	351.946	0.000	100.0#	0#	-4.53#
9 L2 AR-1221-2	277.790	0.000	100.0#	0#	-4.62#
10 L2 AR-1221-3	870.510	0.000	100.0#	0#	-4.69#
11 L3 AR-1232-1	1.178	0.000 E3	100.0#	0#	-4.69#
12 L3 AR-1232-2	652.766	0.000	100.0#	0#	-5.22#
13 L3 AR-1232-3	1.414	0.000 E3	100.0#	0#	-5.51#
14 L3 AR-1232-4	716.970	0.000	100.0#	0#	-5.67#
15 L3 AR-1232-5	558.740	0.000	100.0#	0#	-5.75#
16 L4 AR-1242-1	988.013	0.000	100.0#	0#	-5.48#
17 L4 AR-1242-2	1.439	0.000 E3	100.0#	0#	-5.51#
18 L4 AR-1242-3	912.349	0.000	100.0#	0#	-5.57#
19 L4 AR-1242-4	741.182	0.000	100.0#	0#	-5.66#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110719\
 Data File : P0063547.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Nov 2019 16:18
 Operator : SM/AJ
 Sample : P0110719ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110719

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 04:56:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0110719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 04:54:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
20 L4 AR-1242-5	963.677	0.000	100.0#	0#	-6.40#
21 L5 AR-1248-1	808.655	0.000	100.0#	0#	-5.48#
22 L5 AR-1248-2	1.213	0.000 E3	100.0#	0#	-5.75#
23 L5 AR-1248-3	1.394	0.000 E3	100.0#	0#	-5.96#
24 L5 AR-1248-4	1.705	0.000 E3	100.0#	0#	-6.36#
25 L5 AR-1248-5	1.645	0.000 E3	100.0#	0#	-6.40#
26 L6 AR-1254-1	1.699	0.000 E3	100.0#	0#	-6.33#
27 L6 AR-1254-2	2.764	0.000 E3	100.0#	0#	-6.55#
28 L6 AR-1254-3	3.180	0.000 E3	100.0#	0#	-6.91#
29 L6 AR-1254-4	2.458	0.000 E3	100.0#	0#	-7.20#
30 L6 AR-1254-5	2.791	0.000 E3	100.0#	0#	-7.62#
36 L8 AR-1262-1	3.196	0.000 E3	100.0#	0#	-7.69#
37 L8 AR-1262-2	5.992	0.000 E3	100.0#	0#	-8.23#
38 L8 AR-1262-3	4.034	0.000 E3	100.0#	0#	-8.53#
39 L8 AR-1262-4	3.027	0.000 E3	100.0#	0#	-8.61#
40 L8 AR-1262-5	2.023	0.000 E3	100.0#	0#	-9.24#
41 L9 AR-1268-1	6.944	0.000 E3	100.0#	0#	-8.52#
42 L9 AR-1268-2	6.349	0.000 E3	100.0#	0#	-8.61#
43 L9 AR-1268-3	5.239	0.000 E3	100.0#	0#	-8.83#
44 L9 AR-1268-4	2.237	0.000 E3	100.0#	0#	-9.24#
45 L9 AR-1268-5	16.440	0.000 E3	100.0#	0#	-9.63#

Signal #2

8 L2 AR-1221-1	211.757	0.000	100.0#	0#	-3.79#
9 L2 AR-1221-2	162.843	0.000	100.0#	0#	-3.88#
10 L2 AR-1221-3	537.410	0.000	100.0#	0#	-3.95#
11 L3 AR-1232-1	742.797	0.000	100.0#	0#	-3.95#
12 L3 AR-1232-2	823.146	0.000	100.0#	0#	-4.68#
13 L3 AR-1232-3	434.702	0.000	100.0#	0#	-4.85#
14 L3 AR-1232-4	392.889	0.000	100.0#	0#	-4.93#
15 L3 AR-1232-5	438.266	0.000	100.0#	0#	-5.10#
16 L4 AR-1242-1	603.077	0.000	100.0#	0#	-4.66#
17 L4 AR-1242-2	829.241	0.000	100.0#	0#	-4.68#
18 L4 AR-1242-3	444.161	0.000	100.0#	0#	-4.85#
19 L4 AR-1242-4	450.435	0.000	100.0#	0#	-4.93#
20 L4 AR-1242-5	608.728	0.000	100.0#	0#	-5.45#
21 L5 AR-1248-1	492.631	0.000	100.0#	0#	-4.66#
22 L5 AR-1248-2	683.497	0.000	100.0#	0#	-4.89#
23 L5 AR-1248-3	701.568	0.000	100.0#	0#	-4.93#
24 L5 AR-1248-4	871.865	0.000	100.0#	0#	-5.10#
25 L5 AR-1248-5	913.761	0.000	100.0#	0#	-5.49#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110719\
 Data File : P0063547.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Nov 2019 16:18
 Operator : SM/AJ
 Sample : P0110719ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110719

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 04:56:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0110719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 04:54:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
26 L6 AR-1254-1	1.401	0.000 E3	100.0#	0#	-5.45#
27 L6 AR-1254-2	1.262	0.000 E3	100.0#	0#	-5.60#
28 L6 AR-1254-3	2.156	0.000 E3	100.0#	0#	-6.00#
29 L6 AR-1254-4	1.381	0.000 E3	100.0#	0#	-6.23#
30 L6 AR-1254-5	1.944	0.000 E3	100.0#	0#	-6.65#
36 L8 AR-1262-1	785.942	0.000	100.0#	0#	-6.74#
37 L8 AR-1262-2	3.540	0.000 E3	100.0#	0#	-7.19#
38 L8 AR-1262-3	1.389	0.000 E3	100.0#	0#	-7.47#
39 L8 AR-1262-4	2.619	0.000 E3	100.0#	0#	-7.53#
40 L8 AR-1262-5	1.145	0.000 E3	100.0#	0#	-8.02#
41 L9 AR-1268-1	4.064	0.000 E3	100.0#	0#	-7.47#
42 L9 AR-1268-2	3.684	0.000 E3	100.0#	0#	-7.53#
43 L9 AR-1268-3	3.013	0.000 E3	100.0#	0#	-7.74#
44 L9 AR-1268-4	1.282	0.000 E3	100.0#	0#	-8.02#
45 L9 AR-1268-5	9.105	0.000 E3	100.0#	0#	-8.31#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0