

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110722\
 Data File : PO090232.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2022 18:10
 Operator : YP/AJ
 Sample : N5436-02DL2 50X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 250FERRYBLVD-TFS-02DL2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:02:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
2) SA Decachlor...	10.052	8.491	15839246	6251746	6.464m	6.070
Target Compounds						
21) L5 AR-1248-1	5.481	4.572	6552227	3167608	103.597	139.529 #
22) L5 AR-1248-2	5.758	4.809	7314810	2663685	77.241	81.930
23) L5 AR-1248-3	5.962	4.850	5232006	2635697	53.073	78.192 #
24) L5 AR-1248-4	6.368	5.023	6893548	1843118	64.656	47.659 #
25) L5 AR-1248-5	6.407	5.415	8562305	5909202	81.510	163.199 #
41) L9 AR-1268-1	8.566	7.407	184.0E6	81859861	527.793	519.577
42) L9 AR-1268-2	8.658	7.471	188.8E6	98484996	605.118	701.926
43) L9 AR-1268-3	8.885	7.678	39534730	19838842	147.119	165.313
44) L9 AR-1268-4	9.306	7.966	134.6E6	59128792	1155.563	1088.733
45) L9 AR-1268-5	9.718	8.246	164.4E6	83738014	191.786	239.206



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110722\
Data File : PO090232.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2022 18:10
Operator : YP/AJ
Sample : N5436-02DL2 50X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:02:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 4.307 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.487 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: -0.001 min
Response: 15839246
Conc: 6.46 ng/ml m



#2 Decachlorobiphenyl


R.T.: 8.491 min
Delta R.T.: -0.005 min
Response: 6251746
Conc: 6.07 ng/ml



#21 AR-1248-1


R.T.: 5.481 min
Delta R.T.: 0.000 min
Response: 6552227
Conc: 103.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#21 AR-1248-1

R.T.: 4.572 min
Delta R.T.: -0.001 min
Response: 3167608
Conc: 139.53 ng/ml




#22 AR-1248-2

R.T.: 5.758 min
Delta R.T.: 0.001 min
Response: 7314810
Conc: 77.24 ng/ml



#22 AR-1248-2


R.T.: 4.809 min
Delta R.T.: -0.001 min
Response: 2663685
Conc: 81.93 ng/ml



#23 AR-1248-3


R.T.: 5.962 min
Delta R.T.: 0.000 min
Response: 5232006
Conc: 53.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#23 AR-1248-3

R.T.: 4.850 min
Delta R.T.: -0.001 min
Response: 2635697
Conc: 78.19 ng/ml




#24 AR-1248-4

R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 6893548
Conc: 64.66 ng/ml



#24 AR-1248-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 1843118
Conc: 47.66 ng/ml



#25 AR-1248-5


R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 8562305
Conc: 81.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#25 AR-1248-5

R.T.: 5.415 min
Delta R.T.: -0.003 min
Response: 5909202
Conc: 163.20 ng/ml




#41 AR-1268-1

R.T.: 8.566 min
Delta R.T.: 0.003 min
Response: 184008923
Conc: 527.79 ng/ml



#41 AR-1268-1


R.T.: 7.407 min
Delta R.T.: 0.000 min
Response: 81859861
Conc: 519.58 ng/ml



#42 AR-1268-2


R.T.: 8.658 min
Delta R.T.: 0.003 min
Response: 188776776
Conc: 605.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#42 AR-1268-2

R.T.: 7.471 min
Delta R.T.: 0.000 min
Response: 98484996
Conc: 701.93 ng/ml




#43 AR-1268-3

R.T.: 8.885 min
Delta R.T.: 0.003 min
Response: 39534730
Conc: 147.12 ng/ml



#43 AR-1268-3


R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 19838842
Conc: 165.31 ng/ml



#44 AR-1268-4


R.T.: 9.306 min
Delta R.T.: 0.005 min
Response: 134646859
Conc: 1155.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-02DL2




#44 AR-1268-4

R.T.: 7.966 min
Delta R.T.: 0.000 min
Response: 59128792
Conc: 1088.73 ng/ml



#45 AR-1268-5

R.T.: 9.718 min
Delta R.T.: 0.004 min
Response: 164444684
Conc: 191.79 ng/ml



#45 AR-1268-5

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 83738014
Conc: 239.21 ng/ml