

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110722\
 Data File : PO090249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2022 23:29
 Operator : YP/AJ
 Sample : N5436-09DL 5X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 250FERRYBLVD-TFS-09DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:30:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.301	3.482	11877155	3747349	3.696	3.526
2) SA Decachlor...	10.051	8.488	151.6E6	61166788	61.866	59.389
Target Compounds						
21) L5 AR-1248-1	5.479	4.570	52162588	24680243	824.740	1087.134 #
22) L5 AR-1248-2	5.756	4.807	68709389	25439525	725.542	782.469
23) L5 AR-1248-3	5.960	4.848	55234224	26764706	560.295	794.015 #
24) L5 AR-1248-4	6.366	5.020	58816477	21271221	551.651	550.024
25) L5 AR-1248-5	6.405	5.414	68542171	22064763	652.496	609.379
41) L9 AR-1268-1	8.564	7.404	2046.6E6	797.9E6	5870.381	5064.103
42) L9 AR-1268-2	8.656	7.468	2205.5E6	989.1E6	7069.708	7049.272
43) L9 AR-1268-3	8.882	7.676	416.9E6	174.7E6	1551.370	1455.544
44) L9 AR-1268-4	9.305	7.964	1325.5E6	544.1E6	11375.284	10019.177
45) L9 AR-1268-5	9.715	8.244	2026.9E6	811.7E6	2363.939	2318.737


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110722\
Data File : PO090249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2022 23:29
Operator : YP/AJ
Sample : N5436-09DL 5X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:30:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.301 min
Delta R.T.: -0.006 min
Response: 11877155
Conc: 3.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#1 Tetrachloro-m-xylene

R.T.: 3.482 min
Delta R.T.: -0.004 min
Response: 3747349
Conc: 3.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.051 min
Delta R.T.: -0.003 min
Response: 151606672
Conc: 61.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.488 min
Delta R.T.: -0.008 min
Response: 61166788
Conc: 59.39 ng/ml



#21 AR-1248-1


R.T.: 5.479 min
Delta R.T.: -0.001 min
Response: 52162588
Conc: 824.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#21 AR-1248-1

R.T.: 4.570 min
Delta R.T.: -0.003 min
Response: 24680243
Conc: 1087.13 ng/ml




#22 AR-1248-2

R.T.: 5.756 min
Delta R.T.: 0.000 min
Response: 68709389
Conc: 725.54 ng/ml



#22 AR-1248-2


R.T.: 4.807 min
Delta R.T.: -0.004 min
Response: 25439525
Conc: 782.47 ng/ml



#23 AR-1248-3


R.T.: 5.960 min
Delta R.T.: -0.001 min
Response: 55234224
Conc: 560.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#23 AR-1248-3

R.T.: 4.848 min
Delta R.T.: -0.004 min
Response: 26764706
Conc: 794.02 ng/ml




#24 AR-1248-4

R.T.: 6.366 min
Delta R.T.: -0.001 min
Response: 58816477
Conc: 551.65 ng/ml



#24 AR-1248-4


R.T.: 5.020 min
Delta R.T.: -0.003 min
Response: 21271221
Conc: 550.02 ng/ml



#25 AR-1248-5


R.T.: 6.405 min
Delta R.T.: 0.000 min
Response: 68542171
Conc: 652.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#25 AR-1248-5

R.T.: 5.414 min
Delta R.T.: -0.004 min
Response: 22064763
Conc: 609.38 ng/ml




#41 AR-1268-1

R.T.: 8.564 min
Delta R.T.: 0.000 min
Response: 2046640525
Conc: 5870.38 ng/ml



#41 AR-1268-1


R.T.: 7.404 min
Delta R.T.: -0.002 min
Response: 797854905
Conc: 5064.10 ng/ml



#42 AR-1268-2


R.T.: 8.656 min
Delta R.T.: 0.001 min
Response: 2205513093
Conc: 7069.71 ng/m

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#42 AR-1268-2

R.T.: 7.468 min
Delta R.T.: -0.002 min
Response: 989061460
Conc: 7049.27 ng/ml




#43 AR-1268-3

R.T.: 8.882 min
Delta R.T.: 0.000 min
Response: 416894525
Conc: 1551.37 ng/ml



#43 AR-1268-3


R.T.: 7.676 min
Delta R.T.: -0.003 min
Response: 174676213
Conc: 1455.54 ng/ml



#44 AR-1268-4


R.T.: 9.305 min
Delta R.T.: 0.003 min
Response: 1325454898
Conc: 11375.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
250FERRYBLVD-TFS-09DL




#44 AR-1268-4

R.T.: 7.964 min
Delta R.T.: -0.002 min
Response: 544138956
Conc: 10019.18 ng/ml



#45 AR-1268-5

R.T.: 9.715 min
Delta R.T.: 0.001 min
Response: 2026927180
Conc: 2363.94 ng/ml



#45 AR-1268-5

R.T.: 8.244 min
Delta R.T.: -0.002 min
Response: 811711701
Conc: 2318.74 ng/ml