

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_0\Data\P0110723\  
 Data File : P0099492.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 07 Nov 2023 21:56  
 Operator : YP/AJ  
 Sample : 05258-22DL 400X  
 Misc :  
 ALS Vial : 42 Sample Multiplier: 1

Instrument :  
 ECD\_0  
 ClientSampleId :  
 22-A-22-B-22-CDL

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
 Supervised By :Ankita Jodhani 11/08/2023

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 07 23:58:30 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_0\methods\P0102523.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Oct 25 06:04:36 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|----------|------|------|--------|--------|-------|-------|
|----------|------|------|--------|--------|-------|-------|

System Monitoring Compounds

| Target Compounds |    |           |       |       |          |          |                  |
|------------------|----|-----------|-------|-------|----------|----------|------------------|
| 26)              | L6 | AR-1254-1 | 6.497 | 5.502 | 15559117 | 9266392  | 236.364 261.516  |
| 27)              | L6 | AR-1254-2 | 6.716 | 5.649 | 35291641 | 13414397 | 366.595 419.404  |
| 28)              | L6 | AR-1254-3 | 7.084 | 6.051 | 50867638 | 28938507 | 569.097 591.941  |
| 29)              | L6 | AR-1254-4 | 7.369 | 6.280 | 40191226 | 19782074 | 733.637m 754.844 |
| 30)              | L6 | AR-1254-5 | 7.790 | 6.698 | 62611433 | 38864181 | 929.852 921.940  |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO110723\  
Data File : PO099492.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 07 Nov 2023 21:56  
Operator : YP/AJ  
Sample : 05258-22DL 400X  
Misc :  
ALS Vial : 42 Sample Multiplier: 1



Instrument :  
ECD\_O  
ClientSampleId :  
22-A-22-B-22-CDL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
Supervised By :Ankita Jodhani 11/08/2023

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 07 23:58:30 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO102523.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Oct 25 06:04:36 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min  
Exp R.T. : 4.478 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
22-A-22-B-22-CDL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
Supervised By :Ankita Jodhani 11/08/2023




#1 Tetrachloro-m-xylene

R.T.: 0.000 min  
Exp R.T. : 3.632 min  
Response: 0  
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 10.280 min  
Response: 0  
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 8.630 min  
Response: 0  
Conc: N.D.




#26 AR-1254-1

R.T.: 6.497 min  
Delta R.T.: -0.011 min  
Response: 15559117  
Conc: 236.36 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
22-A-22-B-22-CDL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
Supervised By :Ankita Jodhani 11/08/2023




#26 AR-1254-1

R.T.: 5.502 min  
Delta R.T.: -0.011 min  
Response: 9266392  
Conc: 261.52 ng/ml




#27 AR-1254-2

R.T.: 6.716 min  
Delta R.T.: -0.012 min  
Response: 35291641  
Conc: 366.60 ng/ml



#27 AR-1254-2

R.T.: 5.649 min  
Delta R.T.: -0.011 min  
Response: 13414397  
Conc: 419.40 ng/ml




#28 AR-1254-3

R.T.: 7.084 min  
Delta R.T.: -0.011 min  
Response: 50867638  
Conc: 569.10 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
22-A-22-B-22-CDL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
Supervised By :Ankita Jodhani 11/08/2023




#28 AR-1254-3

R.T.: 6.051 min  
Delta R.T.: -0.012 min  
Response: 28938507  
Conc: 591.94 ng/ml




#29 AR-1254-4

R.T.: 7.369 min  
Delta R.T.: -0.011 min  
Response: 40191226  
Conc: 733.64 ng/ml m



#29 AR-1254-4

R.T.: 6.280 min  
Delta R.T.: -0.012 min  
Response: 19782074  
Conc: 754.84 ng/ml




#30 AR-1254-5

R.T.: 7.790 min  
Delta R.T.: -0.010 min  
Response: 62611433  
Conc: 929.85 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
22-A-22-B-22-CDL

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/08/2023  
Supervised By :Ankita Jodhani 11/08/2023



#30 AR-1254-5

R.T.: 6.698 min  
Delta R.T.: -0.011 min  
Response: 38864181  
Conc: 921.94 ng/ml