

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110724\
 Data File : P0107797.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2024 21:21
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 01:00:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 28 11:34:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.372	3.643	516.6E6	190.7E6	56.679	59.237
2) SA Decachlor...	10.077	8.638	119.0E6	139.6E6	48.484	50.907
Target Compounds						
3) L1 AR-1016-1	5.524	4.723	151.6E6	60798556	561.893	588.759
4) L1 AR-1016-2	5.546	4.742	216.1E6	84431861	544.591	598.944
5) L1 AR-1016-3	5.609	4.918	135.4E6	45388756	539.143	576.397
6) L1 AR-1016-4	5.706	4.960	107.9E6	34161799	559.973	516.755
7) L1 AR-1016-5	6.001	5.173	100.6E6	46956324	550.932	571.705
31) L7 AR-1260-1	7.131	6.204	130.9E6	86215962	506.263	555.678
32) L7 AR-1260-2	7.388	6.391	141.1E6	102.9E6	536.106	583.998
33) L7 AR-1260-3	7.751	6.545	99065517	93097183	551.144	554.982
34) L7 AR-1260-4	7.975	7.015	99946120	78309117	569.481	541.255
35) L7 AR-1260-5	8.290	7.256	161.7E6	186.7E6	568.618	566.716



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110724\
Data File : PO107797.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2024 21:21
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 01:00:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 28 11:34:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.372 min
Delta R.T.: -0.002 min
Response: 516565928
Conc: 56.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: -0.002 min
Response: 190655430
Conc: 59.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.077 min
Delta R.T.: 0.018 min
Response: 118950242
Conc: 48.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.638 min
Delta R.T.: 0.000 min
Response: 139565920
Conc: 50.91 ng/ml



#3 AR-1016-1


R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 151606839
Conc: 561.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.723 min
Delta R.T.: -0.003 min
Response: 60798556
Conc: 588.76 ng/ml




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: 0.001 min
Response: 216139819
Conc: 544.59 ng/ml



#4 AR-1016-2


R.T.: 4.742 min
Delta R.T.: -0.003 min
Response: 84431861
Conc: 598.94 ng/ml



#5 AR-1016-3


R.T.: 5.609 min
Delta R.T.: 0.001 min
Response: 135416368
Conc: 539.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.918 min
Delta R.T.: -0.003 min
Response: 45388756
Conc: 576.40 ng/ml




#6 AR-1016-4

R.T.: 5.706 min
Delta R.T.: 0.002 min
Response: 107857537
Conc: 559.97 ng/ml



#6 AR-1016-4


R.T.: 4.960 min
Delta R.T.: -0.003 min
Response: 34161799
Conc: 516.75 ng/ml



#7 AR-1016-5


R.T.: 6.001 min
Delta R.T.: 0.002 min
Response: 100629774
Conc: 550.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.173 min
Delta R.T.: -0.003 min
Response: 46956324
Conc: 571.70 ng/ml




#31 AR-1260-1

R.T.: 7.131 min
Delta R.T.: 0.005 min
Response: 130913691
Conc: 506.26 ng/ml



#31 AR-1260-1


R.T.: 6.204 min
Delta R.T.: -0.003 min
Response: 86215962
Conc: 555.68 ng/ml



#32 AR-1260-2


R.T.: 7.388 min
Delta R.T.: 0.006 min
Response: 141113787
Conc: 536.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.391 min
Delta R.T.: -0.003 min
Response: 102851239
Conc: 584.00 ng/ml




#33 AR-1260-3

R.T.: 7.751 min
Delta R.T.: 0.008 min
Response: 99065517
Conc: 551.14 ng/ml



#33 AR-1260-3


R.T.: 6.545 min
Delta R.T.: -0.003 min
Response: 93097183
Conc: 554.98 ng/ml



#34 AR-1260-4


R.T.: 7.975 min
Delta R.T.: 0.007 min
Response: 99946120
Conc: 569.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.015 min
Delta R.T.: -0.003 min
Response: 78309117
Conc: 541.25 ng/ml



#35 AR-1260-5

R.T.: 8.290 min
Delta R.T.: 0.008 min
Response: 161711291
Conc: 568.62 ng/ml



#35 AR-1260-5

R.T.: 7.256 min
Delta R.T.: -0.003 min
Response: 186699403
Conc: 566.72 ng/ml