

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110725\
 Data File : PO114975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2025 17:09
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 01:40:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.648	375.2E6	231.8E6	53.316	51.898
2)	SA Decachlor...	9.933	8.642	229.8E6	214.7E6	51.980	54.986
Target Compounds							
16)	L4 AR-1242-1	5.487	4.727	95729095	72411967	514.007	495.221
17)	L4 AR-1242-2	5.509	4.747	152.5E6	102.1E6	510.353	492.959
18)	L4 AR-1242-3	5.570	4.921	95756803	56338390	558.776	503.122
19)	L4 AR-1242-4	5.666	5.005	83489235	53426887	583.092	498.705
20)	L4 AR-1242-5	6.392	5.526	71241579	69911510	505.412	489.173



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110725\
Data File : PO114975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2025 17:09
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 01:40:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 375245920
Conc: 53.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 231816525
Conc: 51.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.933 min
Delta R.T.: 0.003 min
Response: 229766691
Conc: 51.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: 0.002 min
Response: 214656532
Conc: 54.99 ng/ml



#16 AR-1242-1


R.T.: 5.487 min
Delta R.T.: 0.000 min
Response: 95729095
Conc: 514.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.727 min
Delta R.T.: -0.001 min
Response: 72411967
Conc: 495.22 ng/ml




#17 AR-1242-2

R.T.: 5.509 min
Delta R.T.: 0.000 min
Response: 152481173
Conc: 510.35 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 102115400
Conc: 492.96 ng/ml



#18 AR-1242-3


R.T.: 5.570 min
Delta R.T.: 0.000 min
Response: 95756803
Conc: 558.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: -0.001 min
Response: 56338390
Conc: 503.12 ng/ml




#19 AR-1242-4

R.T.: 5.666 min
Delta R.T.: -0.001 min
Response: 83489235
Conc: 583.09 ng/ml



#19 AR-1242-4


R.T.: 5.005 min
Delta R.T.: -0.001 min
Response: 53426887
Conc: 498.71 ng/ml



#20 AR-1242-5

R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 71241579
Conc: 505.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.526 min
Delta R.T.: 0.000 min
Response: 69911510
Conc: 489.17 ng/ml