

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110822\
 Data File : PO090328.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2022 02:38
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 05:07:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.302	3.481	163.0E6	52575536	50.724	49.469
2) SA Decachlor...	10.046	8.484	119.3E6	46954059	48.666	45.590
Target Compounds						
21) L5 AR-1248-1	5.476	4.566	33269689	11322248	526.025	498.731
22) L5 AR-1248-2	5.752	4.803	45635200	16202631	481.888	498.360
23) L5 AR-1248-3	5.957	4.845	50220692	16773556	509.438	497.613
24) L5 AR-1248-4	6.363	5.016	51263217	19650110	480.808	508.106
25) L5 AR-1248-5	6.402	5.410	49752642	18304353	473.626	505.525



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110822\
Data File : PO090328.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2022 02:38
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 05:07:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.302 min
Delta R.T.: -0.005 min
Response: 163022201
Conc: 50.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: -0.006 min
Response: 52575536
Conc: 49.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.008 min
Response: 119259171
Conc: 48.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.484 min
Delta R.T.: -0.013 min
Response: 46954059
Conc: 45.59 ng/ml



#21 AR-1248-1


R.T.: 5.476 min
Delta R.T.: -0.004 min
Response: 33269689
Conc: 526.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.566 min
Delta R.T.: -0.007 min
Response: 11322248
Conc: 498.73 ng/ml




#22 AR-1248-2

R.T.: 5.752 min
Delta R.T.: -0.004 min
Response: 45635200
Conc: 481.89 ng/ml



#22 AR-1248-2


R.T.: 4.803 min
Delta R.T.: -0.007 min
Response: 16202631
Conc: 498.36 ng/ml



#23 AR-1248-3


R.T.: 5.957 min
Delta R.T.: -0.004 min
Response: 50220692
Conc: 509.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.845 min
Delta R.T.: -0.007 min
Response: 16773556
Conc: 497.61 ng/ml




#24 AR-1248-4

R.T.: 6.363 min
Delta R.T.: -0.004 min
Response: 51263217
Conc: 480.81 ng/ml



#24 AR-1248-4


R.T.: 5.016 min
Delta R.T.: -0.007 min
Response: 19650110
Conc: 508.11 ng/ml



#25 AR-1248-5

R.T.: 6.402 min
Delta R.T.: -0.004 min
Response: 49752642
Conc: 473.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.410 min
Delta R.T.: -0.008 min
Response: 18304353
Conc: 505.53 ng/ml