

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0110918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 08 14:50:13 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0051256.D 1000 =P0051254.D 750 =P0051255.D
 250 =P0051257.D 50 =P0051258.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.646	5.617	5.567	5.764	5.594	5.638	E5	1.36
2) SA	Decachlorobiphenyl	3.613	3.691	3.680	3.652	3.917	3.711	E5	3.22
3) L1	AR-1016-1	1.670	1.646	1.651	1.703	1.567	1.648	E4	3.04
4) L1	AR-1016-2	2.528	2.487	2.471	2.550	2.350	2.477	E4	3.14
5) L1	AR-1016-3	1.553	1.510	1.490	1.552	1.532	1.527	E4	1.79
6) L1	AR-1016-4	1.249	1.257	1.210	1.259	1.195	1.234	E4	2.38
7) L1	AR-1016-5	1.154	1.162	1.144	1.164	0.935	1.112	E4	8.91
8) L2	AR-1221-1	5.345					5.345	E3	0.00
9) L2	AR-1221-2	4.055					4.055	E3	0.00
10) L2	AR-1221-3	1.292					1.292	E4	0.00
11) L3	AR-1232-1	9.978					9.978	E3	0.00
12) L3	AR-1232-2	5.241					5.241	E3	0.00
13) L3	AR-1232-3	1.106					1.106	E4	0.00
14) L3	AR-1232-4	5.251					5.251	E3	0.00
15) L3	AR-1232-5	3.748					3.748	E3	0.00
16) L4	AR-1242-1	2.113					2.113	E4	0.00
17) L4	AR-1242-2	2.113					2.113	E4	0.00
18) L4	AR-1242-3	1.271					1.271	E4	0.00
19) L4	AR-1242-4	1.035					1.035	E4	0.00
20) L4	AR-1242-5	1.068					1.068	E4	0.00
21) L5	AR-1248-1	1.156					1.156	E4	0.00
22) L5	AR-1248-2	1.574					1.574	E4	0.00
23) L5	AR-1248-3	1.779					1.779	E4	0.00
24) L5	AR-1248-4	1.995					1.995	E4	0.00
25) L5	AR-1248-5	1.885					1.885	E4	0.00
26) L6	AR-1254-1	1.941					1.941	E4	0.00
27) L6	AR-1254-2	3.022					3.022	E4	0.00
28) L6	AR-1254-3	3.042					3.042	E4	0.00
29) L6	AR-1254-4	2.231					2.231	E4	0.00
30) L6	AR-1254-5	2.228					2.228	E4	0.00
31) L7	AR-1260-1	2.201	2.143	2.144	2.271	2.386	2.229	E4	4.58
32) L7	AR-1260-2	2.834	2.753	2.774	2.911	3.081	2.871	E4	4.63
33) L7	AR-1260-3	1.894	1.841	1.850	2.017	2.172	1.955	E4	7.16
34) L7	AR-1260-4	1.838	1.798	1.814	1.862	2.080	1.878	E4	6.14
35) L7	AR-1260-5	3.718	3.675	3.673	3.792	4.043	3.780	E4	4.09
36) L8	AR-1262-1	3.234					3.234	E4	0.00
37) L8	AR-1262-2	4.920					4.920	E4	0.00
38) L8	AR-1262-3	3.086					3.086	E4	0.00
39) L8	AR-1262-4	2.434					2.434	E4	0.00
40) L8	AR-1262-5	1.569					1.569	E4	0.00
41) L9	AR-1268-1	6.526					6.526	E4	0.00
42) L9	AR-1268-2	5.848					5.848	E4	0.00
43) L9	AR-1268-3	5.208					5.208	E4	0.00
44) L9	AR-1268-4	1.905					1.905	E4	0.00
45) L9	AR-1268-5	1.522					1.522	E5	0.00

Signal #2 Calibration Files

500 =P0051256.D 1000 =P0051254.D 750 =P0051255.D
 250 =P0051257.D 50 =P0051258.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.255	2.227	2.252	2.263	2.160	2.231	E5	1.88
2) SA	Decachlorobiphenyl	2.183	2.091	2.139	2.244	2.424	2.216	E5	5.83
3) L1	AR-1016-1	7.376	6.943	7.278	7.495	7.280	7.275	E3	2.83
4) L1	AR-1016-2	1.256	1.190	1.200	1.364	1.245	1.251	E4	5.53
5) L1	AR-1016-3	6.139	5.980	5.954	6.321	6.028	6.085	E3	2.46

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0110918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 08 14:50:13 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0051256.D 1000 =P0051254.D 750 =P0051255.D
 250 =P0051257.D 50 =P0051258.D

Compound			500	1000	750	250	50	Avg		%RSD

6)	L1	AR-1016-4	5.102	4.858	4.887	5.172	4.987	5.001	E3	2.70
7)	L1	AR-1016-5	6.631	6.263	6.420	6.682	6.412	6.482	E3	2.66
8)	L2	AR-1221-1	2.158					2.158	E3	0.00
9)	L2	AR-1221-2	1.580					1.580	E3	0.00
10)	L2	AR-1221-3	5.348					5.348	E3	0.00
11)	L3	AR-1232-1	4.226					4.226	E3	0.00
12)	L3	AR-1232-2	5.738					5.738	E3	0.00
13)	L3	AR-1232-3	2.565					2.565	E3	0.00
14)	L3	AR-1232-4	2.215					2.215	E3	0.00
15)	L3	AR-1232-5	2.562					2.562	E3	0.00
16)	L4	AR-1242-1	6.314					6.314	E3	0.00
17)	L4	AR-1242-2	1.028					1.028	E4	0.00
18)	L4	AR-1242-3	5.244					5.244	E3	0.00
19)	L4	AR-1242-4	4.959					4.959	E3	0.00
20)	L4	AR-1242-5	6.337					6.337	E3	0.00
21)	L5	AR-1248-1	5.343					5.343	E3	0.00
22)	L5	AR-1248-2	7.838					7.838	E3	0.00
23)	L5	AR-1248-3	7.823					7.823	E3	0.00
24)	L5	AR-1248-4	9.788					9.788	E3	0.00
25)	L5	AR-1248-5	9.831					9.831	E3	0.00
26)	L6	AR-1254-1	1.504					1.504	E4	0.00
27)	L6	AR-1254-2	1.312					1.312	E4	0.00
28)	L6	AR-1254-3	2.100					2.100	E4	0.00
29)	L6	AR-1254-4	1.304					1.304	E4	0.00
30)	L6	AR-1254-5	1.785					1.785	E4	0.00
31)	L7	AR-1260-1	1.397	1.323	1.332	1.446	1.640	1.427	E4	9.03
32)	L7	AR-1260-2	1.829	1.742	1.765	1.916	2.216	1.894	E4	10.16
33)	L7	AR-1260-3	1.596	1.511	1.549	1.647	1.917	1.644	E4	9.77
34)	L7	AR-1260-4	1.094	1.023	1.054	1.138	1.131	1.088	E4	4.56
35)	L7	AR-1260-5	2.904	2.757	2.828	2.875	3.207	2.914	E4	5.94
36)	L8	AR-1262-1	2.316					2.316	E4	0.00
37)	L8	AR-1262-2	3.747					3.747	E4	0.00
38)	L8	AR-1262-3	1.547					1.547	E4	0.00
39)	L8	AR-1262-4	2.607					2.607	E4	0.00
40)	L8	AR-1262-5	1.131					1.131	E4	0.00
41)	L9	AR-1268-1	5.010					5.010	E4	0.00
42)	L9	AR-1268-2	4.359					4.359	E4	0.00
43)	L9	AR-1268-3	3.912					3.912	E4	0.00
44)	L9	AR-1268-4	1.386					1.386	E4	0.00
45)	L9	AR-1268-5	1.003					1.003	E5	0.00

 (#) = Out of Range