

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110918\
 Data File : PO051259.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Nov 2018 12:25
 Operator : SM/SJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 12:45:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 08 12:44:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.349	3.336	27762194	10336691	49.408	47.116
2)	SA Decachlor...	9.995	8.074	16501235	10376397	45.231	47.338
Target Compounds							
3)	L1 AR-1016-1	5.513	4.366	546430	684545	33.144	95.403 #
4)	L1 AR-1016-2	5.513	4.366	546430	684545	21.989	54.723 #
5)	L1 AR-1016-3	5.597	0.000	619888	0	40.763	N.D. #
6)	L1 AR-1016-4	5.696	0.000	383173	0	31.050	N.D. #
8)	L2 AR-1221-1	4.560	3.539	2672538	1078756	500.000	500.000
9)	L2 AR-1221-2	4.644	3.618	2027351	790041	500.000	500.000
10)	L2 AR-1221-3	4.720	3.688	6461491	2674078	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110918\
Data File : PO051259.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Nov 2018 12:25
Operator : SM/SJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 12:45:01 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 08 12:44:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.349 min
Delta R.T.: 0.011 min
Response: 27762194
Conc: 49.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.336 min
Delta R.T.: 0.000 min
Response: 10336691
Conc: 47.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.995 min
Delta R.T.: 0.018 min
Response: 16501235
Conc: 45.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.074 min
Delta R.T.: 0.004 min
Response: 10376397
Conc: 47.34 ng/ml



#3 AR-1016-1


R.T.: 5.513 min
Delta R.T.: 0.013 min
Response: 546430
Conc: 33.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.366 min
Delta R.T.: 0.016 min
Response: 684545
Conc: 95.40 ng/ml




#4 AR-1016-2

R.T.: 5.513 min
Delta R.T.: -0.010 min
Response: 546430
Conc: 21.99 ng/ml



#4 AR-1016-2


R.T.: 4.366 min
Delta R.T.: 0.000 min
Response: 684545
Conc: 54.72 ng/ml



#5 AR-1016-3


R.T.: 5.597 min
Delta R.T.: 0.013 min
Response: 619888
Conc: 40.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.529 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.696 min
Delta R.T.: 0.013 min
Response: 383173
Conc: 31.05 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.573 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.560 min
Delta R.T.: 0.007 min
Response: 2672538
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 3.539 min
Delta R.T.: -0.005 min
Response: 1078756
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.644 min
Delta R.T.: 0.005 min
Response: 2027351
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.618 min
Delta R.T.: -0.005 min
Response: 790041
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.720 min
Delta R.T.: 0.005 min
Response: 6461491
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :



#10 AR-1221-3

R.T.: 3.688 min
Delta R.T.: -0.005 min
Response: 2674078
Conc: 500.00 ng/ml