

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110918\
 Data File : P0051303.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Nov 2018 1:47
 Operator : SM/SJ
 Sample : J5805-06
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-SAMPLES(26-30)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 08:17:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 08 14:50:13 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.336	13115907	5241117	23.265	23.488
2) SA Decachlor...	9.971	8.068	5782604	2445290	15.584	11.033 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110918\
Data File : PO051303.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Nov 2018 1:47
Operator : SM/SJ
Sample : J5805-06
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-SAMPLES(26-30)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 08:17:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 08 14:50:13 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: -0.003 min
Response: 13115907
Conc: 23.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-SAMPLES(26-30)




#1 Tetrachloro-m-xylene

R.T.: 3.336 min
Delta R.T.: 0.000 min
Response: 5241117
Conc: 23.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.971 min
Delta R.T.: -0.006 min
Response: 5782604
Conc: 15.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.068 min
Delta R.T.: -0.002 min
Response: 2445290
Conc: 11.03 ng/ml