

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111022\
 Data File : PO090409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2022 09:37
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 20:46:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.302	3.480	168.1E6	51682434	52.290	48.629
2) SA Decachlor...	10.048	8.483	120.3E6	46781941	49.089	45.423
Target Compounds						
21) L5 AR-1248-1	5.477	4.566	33418939	11325221	528.385	498.862
22) L5 AR-1248-2	5.753	4.803	44745383	16088358	472.492	494.845
23) L5 AR-1248-3	5.958	4.845	50060823	16555943	507.816	491.157
24) L5 AR-1248-4	6.364	5.016	52962895	19442004	496.749	502.725
25) L5 AR-1248-5	6.402	5.411	51164609	18089943	487.068	499.604



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111022\
Data File : PO090409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2022 09:37
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 20:46:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.302 min
Delta R.T.: -0.005 min
Response: 168056473
Conc: 52.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: -0.006 min
Response: 51682434
Conc: 48.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.048 min
Delta R.T.: -0.006 min
Response: 120295359
Conc: 49.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.483 min
Delta R.T.: -0.013 min
Response: 46781941
Conc: 45.42 ng/ml



#21 AR-1248-1


R.T.: 5.477 min
Delta R.T.: -0.004 min
Response: 33418939
Conc: 528.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.566 min
Delta R.T.: -0.007 min
Response: 11325221
Conc: 498.86 ng/ml




#22 AR-1248-2

R.T.: 5.753 min
Delta R.T.: -0.003 min
Response: 44745383
Conc: 472.49 ng/ml



#22 AR-1248-2


R.T.: 4.803 min
Delta R.T.: -0.007 min
Response: 16088358
Conc: 494.85 ng/ml



#23 AR-1248-3


R.T.: 5.958 min
Delta R.T.: -0.003 min
Response: 50060823
Conc: 507.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.845 min
Delta R.T.: -0.007 min
Response: 16555943
Conc: 491.16 ng/ml




#24 AR-1248-4

R.T.: 6.364 min
Delta R.T.: -0.003 min
Response: 52962895
Conc: 496.75 ng/ml



#24 AR-1248-4


R.T.: 5.016 min
Delta R.T.: -0.007 min
Response: 19442004
Conc: 502.72 ng/ml



#25 AR-1248-5

R.T.: 6.402 min
Delta R.T.: -0.003 min
Response: 51164609
Conc: 487.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.411 min
Delta R.T.: -0.008 min
Response: 18089943
Conc: 499.60 ng/ml