

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111022\
 Data File : PO090414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2022 11:02
 Operator : YP/AJ
 Sample : N5528-01DL 4X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 SAMPLE-3DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/11/2022
 Supervised By :Ankita Jodhani 11/11/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 20:54:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.303	3.482	14154387	4049551	4.404	3.810
2) SA Decachlor...	10.050	8.483	8067359	6252747	3.292	6.071 #
Target Compounds						
26) L6 AR-1254-1	6.339	5.372	44221790	21620877	381.561	367.044
27) L6 AR-1254-2	6.554	5.520	57961555	21605208	333.240m	403.230
28) L6 AR-1254-3	6.926	5.925	83531726	39572856	481.159	477.457
29) L6 AR-1254-4	7.214	6.155	60693185	20725992	476.150	417.320
30) L6 AR-1254-5	7.637	6.574	83069769	42066001	589.735	561.813

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111022\
Data File : PO090414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2022 11:02
Operator : YP/AJ
Sample : N5528-01DL 4X
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SAMPLE-3DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/11/2022
Supervised By :Ankita Jodhani 11/11/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 20:54:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.303 min
Delta R.T.: -0.004 min
Response: 14154387
Conc: 4.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
SAMPLE-3DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/11/2022
Supervised By :Ankita Jodhani 11/11/2022




#1 Tetrachloro-m-xylene

R.T.: 3.482 min
Delta R.T.: -0.005 min
Response: 4049551
Conc: 3.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.050 min
Delta R.T.: -0.004 min
Response: 8067359
Conc: 3.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.483 min
Delta R.T.: -0.014 min
Response: 6252747
Conc: 6.07 ng/ml




#26 AR-1254-1

R.T.: 6.339 min
Delta R.T.: 0.000 min
Response: 44221790
Conc: 381.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
SAMPLE-3DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/11/2022
Supervised By :Ankita Jodhani 11/11/2022




#26 AR-1254-1

R.T.: 5.372 min
Delta R.T.: -0.004 min
Response: 21620877
Conc: 367.04 ng/ml




#27 AR-1254-2

R.T.: 6.554 min
Delta R.T.: -0.006 min
Response: 57961555
Conc: 333.24 ng/ml m



#27 AR-1254-2

R.T.: 5.520 min
Delta R.T.: -0.005 min
Response: 21605208
Conc: 403.23 ng/ml




#28 AR-1254-3

R.T.: 6.926 min
Delta R.T.: -0.001 min
Response: 83531726
Conc: 481.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
SAMPLE-3DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/11/2022
Supervised By :Ankita Jodhani 11/11/2022




#28 AR-1254-3

R.T.: 5.925 min
Delta R.T.: -0.005 min
Response: 39572856
Conc: 477.46 ng/ml




#29 AR-1254-4

R.T.: 7.214 min
Delta R.T.: 0.000 min
Response: 60693185
Conc: 476.15 ng/ml



#29 AR-1254-4

R.T.: 6.155 min
Delta R.T.: -0.005 min
Response: 20725992
Conc: 417.32 ng/ml



#30 AR-1254-5

R.T.: 7.637 min
Delta R.T.: 0.000 min
Response: 83069769
Conc: 589.73 ng/ml

Instrument :

ECD_O

ClientSampleId :


SAMPLE-3DL

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/11/2022

Supervised By :Ankita Jodhani 11/11/2022



#30 AR-1254-5

R.T.: 6.574 min
Delta R.T.: -0.005 min
Response: 42066001
Conc: 561.81 ng/ml