

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111022\
 Data File : PO090445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2022 21:21
 Operator : YP/AJ
 Sample : N5531-06MS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-34MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 03:47:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.302	3.481	72242359	21922258	22.478	20.627
2) SA Decachlor...	10.050	8.485	43466330	18606336	17.737	18.066
Target Compounds						
3) L1 AR-1016-1	5.478	4.567	58114457	20047019	546.355	518.992
4) L1 AR-1016-2	5.500	4.585	82795388	28399135	543.515	530.893
5) L1 AR-1016-3	5.562	4.761	51930625	15675934	549.599	539.938
6) L1 AR-1016-4	5.661	4.804	41211188	13130908	549.569	528.983
7) L1 AR-1016-5	5.958	5.017	42099327	16684154	545.462	543.118
31) L7 AR-1260-1	7.093	6.055	73892140	31702414	527.103	526.607
32) L7 AR-1260-2	7.352	6.245	86234826	36888705	533.038	510.624
33) L7 AR-1260-3	7.714	6.398	53421422	33711748	508.507	505.926
34) L7 AR-1260-4	7.940	6.872	63401304	25661870	513.197	504.418
35) L7 AR-1260-5	8.256	7.116	113.5E6	57473916	498.661	494.818



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111022\
Data File : PO090445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2022 21:21
Operator : YP/AJ
Sample : N5531-06MS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-34MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 03:47:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.302 min
Delta R.T.: -0.005 min
Response: 72242359
Conc: 22.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: -0.006 min
Response: 21922258
Conc: 20.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.050 min
Delta R.T.: -0.004 min
Response: 43466330
Conc: 17.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.485 min
Delta R.T.: -0.011 min
Response: 18606336
Conc: 18.07 ng/ml



#3 AR-1016-1


R.T.: 5.478 min
Delta R.T.: -0.004 min
Response: 58114457
Conc: 546.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#3 AR-1016-1

R.T.: 4.567 min
Delta R.T.: -0.007 min
Response: 20047019
Conc: 518.99 ng/ml




#4 AR-1016-2

R.T.: 5.500 min
Delta R.T.: -0.004 min
Response: 82795388
Conc: 543.52 ng/ml



#4 AR-1016-2


R.T.: 4.585 min
Delta R.T.: -0.007 min
Response: 28399135
Conc: 530.89 ng/ml



#5 AR-1016-3


R.T.: 5.562 min
Delta R.T.: -0.004 min
Response: 51930625
Conc: 549.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#5 AR-1016-3

R.T.: 4.761 min
Delta R.T.: -0.008 min
Response: 15675934
Conc: 539.94 ng/ml




#6 AR-1016-4

R.T.: 5.661 min
Delta R.T.: -0.004 min
Response: 41211188
Conc: 549.57 ng/ml



#6 AR-1016-4


R.T.: 4.804 min
Delta R.T.: -0.007 min
Response: 13130908
Conc: 528.98 ng/ml



#7 AR-1016-5


R.T.: 5.958 min
Delta R.T.: -0.004 min
Response: 42099327
Conc: 545.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#7 AR-1016-5

R.T.: 5.017 min
Delta R.T.: -0.008 min
Response: 16684154
Conc: 543.12 ng/ml




#31 AR-1260-1

R.T.: 7.093 min
Delta R.T.: -0.003 min
Response: 73892140
Conc: 527.10 ng/ml



#31 AR-1260-1


R.T.: 6.055 min
Delta R.T.: -0.009 min
Response: 31702414
Conc: 526.61 ng/ml



#32 AR-1260-2


R.T.: 7.352 min
Delta R.T.: -0.002 min
Response: 86234826
Conc: 533.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#32 AR-1260-2

R.T.: 6.245 min
Delta R.T.: -0.008 min
Response: 36888705
Conc: 510.62 ng/ml




#33 AR-1260-3

R.T.: 7.714 min
Delta R.T.: -0.002 min
Response: 53421422
Conc: 508.51 ng/ml



#33 AR-1260-3


R.T.: 6.398 min
Delta R.T.: -0.008 min
Response: 33711748
Conc: 505.93 ng/ml



#34 AR-1260-4


R.T.: 7.940 min
Delta R.T.: -0.002 min
Response: 63401304
Conc: 513.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-34MS




#34 AR-1260-4

R.T.: 6.872 min
Delta R.T.: -0.009 min
Response: 25661870
Conc: 504.42 ng/ml



#35 AR-1260-5

R.T.: 8.256 min
Delta R.T.: -0.002 min
Response: 113474120
Conc: 498.66 ng/ml



#35 AR-1260-5

R.T.: 7.116 min
Delta R.T.: -0.009 min
Response: 57473916
Conc: 494.82 ng/ml