

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111023\
 Data File : PO099583.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2023 14:37
 Operator : YP/AJ
 Sample : PB157045BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB157045BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 22:29:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 10 11:13:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.494	3.723	39960379	27611058	23.213	22.775
2)	SA Decachlor...	10.381	8.839	14031454	17488866	21.982	23.810
Target Compounds							
3)	L1 AR-1016-1	5.682	4.831	1895673	1332231	48.779	49.544
4)	L1 AR-1016-2	5.704	4.851	2826356	1641618	44.130	45.038
5)	L1 AR-1016-3	5.767	5.030	1916519	843593	44.228	42.093
6)	L1 AR-1016-4	5.867	5.073	1374852	806626	44.241	48.538
7)	L1 AR-1016-5	6.165	5.289	1472185	854884	48.408	41.384
31)	L7 AR-1260-1	7.305	6.338	2016208	2070923	40.566	48.017
32)	L7 AR-1260-2	7.565	6.527	2358147	2280942	46.167	48.680
33)	L7 AR-1260-3	7.930	6.683	1468385	2085119	40.956	45.685
34)	L7 AR-1260-4	8.158	7.162	1703676	1706432	43.157	46.436
35)	L7 AR-1260-5	8.488	7.403	2647138	3112331	44.186	42.699



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111023\
Data File : PO099583.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2023 14:37
Operator : YP/AJ
Sample : PB157045BSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB157045BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 22:29:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110923.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 10 11:13:12 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.494 min
Delta R.T.: 0.000 min
Response: 39960379
Conc: 23.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#1 Tetrachloro-m-xylene

R.T.: 3.723 min
Delta R.T.: -0.001 min
Response: 27611058
Conc: 22.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.381 min
Delta R.T.: 0.000 min
Response: 14031454
Conc: 21.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.839 min
Delta R.T.: -0.005 min
Response: 17488866
Conc: 23.81 ng/ml



#3 AR-1016-1


R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 1895673
Conc: 48.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#3 AR-1016-1

R.T.: 4.831 min
Delta R.T.: -0.002 min
Response: 1332231
Conc: 49.54 ng/ml




#4 AR-1016-2

R.T.: 5.704 min
Delta R.T.: -0.001 min
Response: 2826356
Conc: 44.13 ng/ml



#4 AR-1016-2


R.T.: 4.851 min
Delta R.T.: -0.002 min
Response: 1641618
Conc: 45.04 ng/ml



#5 AR-1016-3


R.T.: 5.767 min
Delta R.T.: 0.000 min
Response: 1916519
Conc: 44.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#5 AR-1016-3

R.T.: 5.030 min
Delta R.T.: -0.002 min
Response: 843593
Conc: 42.09 ng/ml




#6 AR-1016-4

R.T.: 5.867 min
Delta R.T.: 0.000 min
Response: 1374852
Conc: 44.24 ng/ml



#6 AR-1016-4


R.T.: 5.073 min
Delta R.T.: -0.001 min
Response: 806626
Conc: 48.54 ng/ml



#7 AR-1016-5


R.T.: 6.165 min
Delta R.T.: -0.001 min
Response: 1472185
Conc: 48.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#7 AR-1016-5

R.T.: 5.289 min
Delta R.T.: -0.003 min
Response: 854884
Conc: 41.38 ng/ml




#31 AR-1260-1

R.T.: 7.305 min
Delta R.T.: 0.000 min
Response: 2016208
Conc: 40.57 ng/ml



#31 AR-1260-1


R.T.: 6.338 min
Delta R.T.: -0.003 min
Response: 2070923
Conc: 48.02 ng/ml



#32 AR-1260-2


R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 2358147
Conc: 46.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#32 AR-1260-2

R.T.: 6.527 min
Delta R.T.: -0.003 min
Response: 2280942
Conc: 48.68 ng/ml




#33 AR-1260-3

R.T.: 7.930 min
Delta R.T.: 0.000 min
Response: 1468385
Conc: 40.96 ng/ml



#33 AR-1260-3


R.T.: 6.683 min
Delta R.T.: -0.003 min
Response: 2085119
Conc: 45.68 ng/ml



#34 AR-1260-4


R.T.: 8.158 min
Delta R.T.: 0.000 min
Response: 1703676
Conc: 43.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB157045BSD




#34 AR-1260-4

R.T.: 7.162 min
Delta R.T.: -0.003 min
Response: 1706432
Conc: 46.44 ng/ml



#35 AR-1260-5

R.T.: 8.488 min
Delta R.T.: 0.000 min
Response: 2647138
Conc: 44.19 ng/ml



#35 AR-1260-5

R.T.: 7.403 min
Delta R.T.: -0.003 min
Response: 3112331
Conc: 42.70 ng/ml