

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111119\
 Data File : P0063632.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Nov 2019 19:07
 Operator : SM/AJ
 Sample : K5763-07
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-34-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 00:39:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 04:54:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.320	3.575	770207	559308	23.276	26.527
2) SA Decachlor...	9.947	8.550	733083	461823	13.360	15.580

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111119\
Data File : P0063632.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Nov 2019 19:07
Operator : SM/AJ
Sample : K5763-07
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-34-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 00:39:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110719.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 04:54:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.320 min
Delta R.T.: -0.004 min
Response: 770207
Conc: 23.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-34-S




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.002 min
Response: 559308
Conc: 26.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.947 min
Delta R.T.: -0.005 min
Response: 733083
Conc: 13.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.550 min
Delta R.T.: -0.004 min
Response: 461823
Conc: 15.58 ng/ml