

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111122\
 Data File : PO090535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2022 05:44
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 07:16:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.305	3.483	187.2E6	62809265	58.255	59.098
2) SA Decachlor...	10.061	8.493	130.9E6	58511855	53.435	56.812
Target Compounds						
16) L4 AR-1242-1	5.482	4.570	47112560	17341988	569.398	582.499
17) L4 AR-1242-2	5.504	4.588	67153450	23917283	568.608	588.010
18) L4 AR-1242-3	5.566	4.764	42221395	13222133	567.212	597.952
19) L4 AR-1242-4	5.664	4.849	33968849	13928828	583.062	587.376
20) L4 AR-1242-5	6.407	5.374	36027650	17477894	586.108	611.008



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111122\
Data File : PO090535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2022 05:44
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 07:16:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.305 min
Delta R.T.: -0.002 min
Response: 187225876
Conc: 58.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.483 min
Delta R.T.: -0.004 min
Response: 62809265
Conc: 59.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.007 min
Response: 130946404
Conc: 53.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.493 min
Delta R.T.: -0.004 min
Response: 58511855
Conc: 56.81 ng/ml



#16 AR-1242-1


R.T.: 5.482 min
Delta R.T.: 0.000 min
Response: 47112560
Conc: 569.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.570 min
Delta R.T.: -0.002 min
Response: 17341988
Conc: 582.50 ng/ml




#17 AR-1242-2

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 67153450
Conc: 568.61 ng/ml



#17 AR-1242-2


R.T.: 4.588 min
Delta R.T.: -0.002 min
Response: 23917283
Conc: 588.01 ng/ml



#18 AR-1242-3


R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 42221395
Conc: 567.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.764 min
Delta R.T.: -0.002 min
Response: 13222133
Conc: 597.95 ng/ml




#19 AR-1242-4

R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 33968849
Conc: 583.06 ng/ml



#19 AR-1242-4


R.T.: 4.849 min
Delta R.T.: -0.002 min
Response: 13928828
Conc: 587.38 ng/ml



#20 AR-1242-5

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 36027650
Conc: 586.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.374 min
Delta R.T.: -0.003 min
Response: 17477894
Conc: 611.01 ng/ml