

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0111221\
 Data File : P0082790.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Nov 2021 00:36
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 07:48:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0111221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 12 07:46:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	89.170	94.861 E3	-6.4	104	0.00
2 SA Decachlorobiphenyl	57.914	64.604 E3	-11.6	108	0.00
16 L4 AR-1242-1	2.142	2.185 E3	-2.0	100	0.00
17 L4 AR-1242-2	3.025	3.078 E3	-1.8	101	0.00
18 L4 AR-1242-3	1.900	1.895 E3	0.3	100	0.00
19 L4 AR-1242-4	1.502	1.538 E3	-2.4	101	0.00
20 L4 AR-1242-5	1.554	1.564 E3	-0.6	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	32.449	34.597 E3	-6.6	106	0.00
2 SA Decachlorobiphenyl	28.353	32.493 E3	-14.6	112	0.00
16 L4 AR-1242-1	1.047	1.070 E3	-2.2	103	0.00
17 L4 AR-1242-2	1.441	1.448 E3	-0.5	102	0.00
18 L4 AR-1242-3	778.673	803.264	-3.2	102	0.00
19 L4 AR-1242-4	767.163	775.706	-1.1	102	0.00
20 L4 AR-1242-5	1.004	1.035 E3	-3.1	103	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	2.944	0.000 E3	100.0#	0#	-6.30#
4 L1 AR-1016-2	4.156	0.000 E3	100.0#	0#	-6.32#
5 L1 AR-1016-3	2.544	0.000 E3	100.0#	0#	-6.39#
6 L1 AR-1016-4	2.050	0.000 E3	100.0#	0#	-6.49#
7 L1 AR-1016-5	2.040	0.000 E3	100.0#	0#	-6.81#
8 L2 AR-1221-1	978.102	0.000	100.0#	0#	-5.22#
9 L2 AR-1221-2	688.709	0.000	100.0#	0#	-5.32#
10 L2 AR-1221-3	2.192	0.000 E3	100.0#	0#	-5.40#
11 L3 AR-1232-1	1.755	0.000 E3	100.0#	0#	-5.41#
12 L3 AR-1232-2	908.413	0.000	100.0#	0#	-6.00#
13 L3 AR-1232-3	1.620	0.000 E3	100.0#	0#	-6.32#
14 L3 AR-1232-4	801.383	0.000	100.0#	0#	-6.50#
15 L3 AR-1232-5	571.563	0.000	100.0#	0#	-6.60#
21 L5 AR-1248-1	1.597	0.000 E3	100.0#	0#	-6.30#
22 L5 AR-1248-2	2.140	0.000 E3	100.0#	0#	-6.59#
23 L5 AR-1248-3	2.516	0.000 E3	100.0#	0#	-6.81#
24 L5 AR-1248-4	2.627	0.000 E3	100.0#	0#	-7.24#
25 L5 AR-1248-5	2.589	0.000 E3	100.0#	0#	-7.28#
26 L6 AR-1254-1	2.795	0.000 E3	100.0#	0#	-7.21#
27 L6 AR-1254-2	4.241	0.000 E3	100.0#	0#	-7.44#
28 L6 AR-1254-3	4.255	0.000 E3	100.0#	0#	-7.83#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111221\
 Data File : PO082790.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Nov 2021 00:36
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 07:48:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO111221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 12 07:46:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	2.984	0.000 E3	100.0#	0#	-8.13#
30 L6 AR-1254-5	3.096	0.000 E3	100.0#	0#	-8.56#
31 L7 AR-1260-1	2.975	0.000 E3	100.0#	0#	-8.00#
32 L7 AR-1260-2	3.673	0.000 E3	100.0#	0#	-8.26#
33 L7 AR-1260-3	2.779	0.000 E3	100.0#	0#	-8.63#
34 L7 AR-1260-4	3.150	0.000 E3	100.0#	0#	-8.86#
35 L7 AR-1260-5	6.194	0.000 E3	100.0#	0#	-9.20#
36 L8 AR-1262-1	3.612	0.000 E3	100.0#	0#	-8.63#
37 L8 AR-1262-2	6.307	0.000 E3	100.0#	0#	-9.20#
38 L8 AR-1262-3	2.892	0.000 E3	100.0#	0#	-9.51#
39 L8 AR-1262-4	1.775	0.000 E3	100.0#	0#	-9.61#
40 L8 AR-1262-5	2.400	0.000 E3	100.0#	0#	-10.29#
41 L9 AR-1268-1	8.088	0.000 E3	100.0#	0#	-9.51#
42 L9 AR-1268-2	7.640	0.000 E3	100.0#	0#	-9.61#
43 L9 AR-1268-3	6.481	0.000 E3	100.0#	0#	-9.84#
44 L9 AR-1268-4	2.862	0.000 E3	100.0#	0#	-10.29#
45 L9 AR-1268-5	21.687	0.000 E3	100.0#	0#	-10.72#

Signal #2

3 L1 AR-1016-1	1.399	0.000 E3	100.0#	0#	-5.22#
4 L1 AR-1016-2	1.935	0.000 E3	100.0#	0#	-5.24#
5 L1 AR-1016-3	1.040	0.000 E3	100.0#	0#	-5.43#
6 L1 AR-1016-4	874.700	0.000	100.0#	0#	-5.48#
7 L1 AR-1016-5	1.070	0.000 E3	100.0#	0#	-5.71#
8 L2 AR-1221-1	412.173	0.000	100.0#	0#	-4.22#
9 L2 AR-1221-2	315.095	0.000	100.0#	0#	-4.32#
10 L2 AR-1221-3	988.999	0.000	100.0#	0#	-4.41#
11 L3 AR-1232-1	810.011	0.000	100.0#	0#	-4.41#
12 L3 AR-1232-2	796.298	0.000	100.0#	0#	-5.24#
13 L3 AR-1232-3	422.212	0.000	100.0#	0#	-5.43#
14 L3 AR-1232-4	365.295	0.000	100.0#	0#	-5.53#
15 L3 AR-1232-5	400.952	0.000	100.0#	0#	-5.71#
21 L5 AR-1248-1	783.689	0.000	100.0#	0#	-5.22#
22 L5 AR-1248-2	1.102	0.000 E3	100.0#	0#	-5.48#
23 L5 AR-1248-3	1.106	0.000 E3	100.0#	0#	-5.53#
24 L5 AR-1248-4	1.305	0.000 E3	100.0#	0#	-5.71#
25 L5 AR-1248-5	1.349	0.000 E3	100.0#	0#	-6.14#
26 L6 AR-1254-1	2.062	0.000 E3	100.0#	0#	-6.09#
27 L6 AR-1254-2	1.805	0.000 E3	100.0#	0#	-6.25#
28 L6 AR-1254-3	2.772	0.000 E3	100.0#	0#	-6.67#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0111221\
 Data File : P0082790.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Nov 2021 00:36
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 07:48:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0111221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 12 07:46:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
29 L6 AR-1254-4	2.004	0.000 E3	100.0#	0#	-6.91#
30 L6 AR-1254-5	2.528	0.000 E3	100.0#	0#	-7.35#
31 L7 AR-1260-1	1.968	0.000 E3	100.0#	0#	-6.81#
32 L7 AR-1260-2	2.543	0.000 E3	100.0#	0#	-7.01#
33 L7 AR-1260-3	2.281	0.000 E3	100.0#	0#	-7.16#
34 L7 AR-1260-4	1.741	0.000 E3	100.0#	0#	-7.65#
35 L7 AR-1260-5	4.380	0.000 E3	100.0#	0#	-7.90#
36 L8 AR-1262-1	1.314	0.000 E3	100.0#	0#	-7.34#
37 L8 AR-1262-2	4.858	0.000 E3	100.0#	0#	-7.90#
38 L8 AR-1262-3	1.965	0.000 E3	100.0#	0#	-8.19#
39 L8 AR-1262-4	3.321	0.000 E3	100.0#	0#	-8.25#
40 L8 AR-1262-5	1.471	0.000 E3	100.0#	0#	-8.75#
41 L9 AR-1268-1	6.130	0.000 E3	100.0#	0#	-8.18#
42 L9 AR-1268-2	5.196	0.000 E3	100.0#	0#	-8.25#
43 L9 AR-1268-3	4.312	0.000 E3	100.0#	0#	-8.46#
44 L9 AR-1268-4	1.731	0.000 E3	100.0#	0#	-8.75#
45 L9 AR-1268-5	11.851	0.000 E3	100.0#	0#	-9.03#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0