

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111221\
 Data File : PO082781.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Nov 2021 22:01
 Operator : AJ\MA
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 06:49:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 12 06:48:43 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.963	3.960	2236807	853596	24.684	25.439
2)	SA Decachlor...	11.061	9.272	1583531	786697	25.521	25.719
Target Compounds							
26)	L6 AR-1254-1	7.212	6.092	718285	537086	259.383	262.367
27)	L6 AR-1254-2	7.444	6.252	1088230	469600	259.081	262.900
28)	L6 AR-1254-3	7.829	6.672	1105221	721637	259.172	262.132
29)	L6 AR-1254-4	8.126	6.914	768231	519464	255.961	260.632
30)	L6 AR-1254-5	8.556	7.343	790699	667934	254.384	261.858



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111221\
Data File : PO082781.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Nov 2021 22:01
Operator : AJ\MA
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 06:49:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 12 06:48:43 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.963 min
Delta R.T.: 0.000 min
Response: 2236807
Conc: 24.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.960 min
Delta R.T.: 0.000 min
Response: 853596
Conc: 25.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.061 min
Delta R.T.: 0.000 min
Response: 1583531
Conc: 25.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.272 min
Delta R.T.: 0.000 min
Response: 786697
Conc: 25.72 ng/ml



#26 AR-1254-1


R.T.: 7.212 min
Delta R.T.: 0.001 min
Response: 718285
Conc: 259.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 6.092 min
Delta R.T.: 0.000 min
Response: 537086
Conc: 262.37 ng/ml




#27 AR-1254-2

R.T.: 7.444 min
Delta R.T.: 0.001 min
Response: 1088230
Conc: 259.08 ng/ml



#27 AR-1254-2


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 469600
Conc: 262.90 ng/ml



#28 AR-1254-3


R.T.: 7.829 min
Delta R.T.: 0.001 min
Response: 1105221
Conc: 259.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.672 min
Delta R.T.: 0.000 min
Response: 721637
Conc: 262.13 ng/ml




#29 AR-1254-4

R.T.: 8.126 min
Delta R.T.: 0.000 min
Response: 768231
Conc: 255.96 ng/ml



#29 AR-1254-4


R.T.: 6.914 min
Delta R.T.: 0.000 min
Response: 519464
Conc: 260.63 ng/ml



#30 AR-1254-5

R.T.: 8.556 min
Delta R.T.: 0.000 min
Response: 790699
Conc: 254.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 667934
Conc: 261.86 ng/ml