

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111224\
 Data File : P0107893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 17:09
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 02:46:57 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 12 01:21:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.373	3.640	473.0E6	180.0E6	52.692	53.587
2) SA Decachlor...	10.085	8.636	142.4E6	163.9E6	48.026	51.894
Target Compounds						
16) L4 AR-1242-1	5.526	4.721	117.0E6	46637474	500.068	515.853
17) L4 AR-1242-2	5.549	4.741	168.3E6	65384505	497.748	516.079
18) L4 AR-1242-3	5.611	4.916	107.9E6	38748125	498.273	567.714
19) L4 AR-1242-4	5.708	5.000	86403076	35353889	505.609	544.903
20) L4 AR-1242-5	6.445	5.521	94358436	46314200	565.875	561.884

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111224\
Data File : PO107893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 17:09
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 02:46:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 12 01:21:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

