

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0111318.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 15 07:59:46 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0051462.D 1000 =P0051460.D 750 =P0051461.D
 250 =P0051463.D 50 =P0051464.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.447	2.827	3.174	3.335	3.185	3.194	E6	7.33
2) SA	Decachlorobiphenyl	2.080	1.831	1.960	1.909	1.816	1.919	E6	5.60
3) L1	AR-1016-1	8.036	6.553	7.432	7.786	7.522	7.466	E4	7.53
4) L1	AR-1016-2	1.211	0.996	1.113	1.182	1.152	1.131	E5	7.42
5) L1	AR-1016-3	7.014	5.766	6.491	6.878	6.698	6.570	E4	7.46
6) L1	AR-1016-4	5.817	4.796	5.391	5.676	5.507	5.437	E4	7.24
7) L1	AR-1016-5	5.257	4.312	4.776	5.018	3.785	4.630	E4	12.69
8) L2	AR-1221-1	3.207	2.709	2.974	2.985	3.067	2.988	E4	6.08
9) L2	AR-1221-2	2.274	1.943	2.125	2.149	2.342	2.166	E4	7.09
10) L2	AR-1221-3	7.349	6.325	6.908	7.080	7.622	7.057	E4	6.95
11) L3	AR-1232-1	5.180					5.180	E4	0.00
12) L3	AR-1232-2	2.393					2.393	E4	0.00
13) L3	AR-1232-3	4.800					4.800	E4	0.00
14) L3	AR-1232-4	2.351					2.351	E4	0.00
15) L3	AR-1232-5	1.554					1.554	E4	0.00
16) L4	AR-1242-1	6.154					6.154	E4	0.00
17) L4	AR-1242-2	9.284					9.284	E4	0.00
18) L4	AR-1242-3	5.368					5.368	E4	0.00
19) L4	AR-1242-4	4.459					4.459	E4	0.00
20) L4	AR-1242-5	4.264					4.264	E4	0.00
21) L5	AR-1248-1	6.526					6.526	E4	0.00
22) L5	AR-1248-2	6.724					6.724	E4	0.00
23) L5	AR-1248-3	7.447					7.447	E4	0.00
24) L5	AR-1248-4	8.317					8.317	E4	0.00
25) L5	AR-1248-5	7.879					7.879	E4	0.00
26) L6	AR-1254-1	7.529					7.529	E4	0.00
27) L6	AR-1254-2	1.174					1.174	E5	0.00
28) L6	AR-1254-3	1.224					1.224	E5	0.00
29) L6	AR-1254-4	9.844					9.844	E4	0.00
30) L6	AR-1254-5	9.880					9.880	E4	0.00
31) L7	AR-1260-1	9.581	7.910	8.785	9.463	8.797	8.907	E4	7.50
32) L7	AR-1260-2	1.359	1.122	1.248	1.308	1.313	1.270	E5	7.21
33) L7	AR-1260-3	9.288	7.769	8.578	9.004	8.817	8.691	E4	6.64
34) L7	AR-1260-4	9.507	8.013	8.775	9.035	8.612	8.788	E4	6.26
35) L7	AR-1260-5	2.234	1.936	2.092	2.095	1.938	2.059	E5	6.09
36) L8	AR-1262-1	1.434					1.434	E5	0.00
37) L8	AR-1262-2	2.497					2.497	E5	0.00
38) L8	AR-1262-3	1.575					1.575	E5	0.00
39) L8	AR-1262-4	1.216					1.216	E5	0.00
40) L8	AR-1262-5	8.028					8.028	E4	0.00
41) L9	AR-1268-1	3.571					3.571	E5	0.00
42) L9	AR-1268-2	3.126					3.126	E5	0.00
43) L9	AR-1268-3	2.791					2.791	E5	0.00
44) L9	AR-1268-4	1.012					1.012	E5	0.00
45) L9	AR-1268-5	8.193					8.193	E5	0.00

Signal #2 Calibration Files

500 =P0051462.D 1000 =P0051460.D 750 =P0051461.D
 250 =P0051463.D 50 =P0051464.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.193	1.007	1.103	1.148	1.055	1.101	E6	6.68
2) SA	Decachlorobiphenyl	8.790	7.317	8.023	8.613	8.704	8.289	E5	7.49
3) L1	AR-1016-1	2.601	2.106	2.356	2.625	2.498	2.437	E4	8.75
4) L1	AR-1016-2	5.715	4.662	5.177	5.654	5.276	5.297	E4	8.01
5) L1	AR-1016-3	2.442	1.974	2.207	2.469	2.351	2.289	E4	8.89

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0111318.M
 Title : GC EXTRACTABLES
 Last Update : Thu Nov 15 07:59:46 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0051462.D 1000 =P0051460.D 750 =P0051461.D
 250 =P0051463.D 50 =P0051464.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.889	1.518	1.706	1.934	1.771	1.764	E4	9.33
7)	L1 AR-1016-5	2.396	1.887	2.148	2.443	2.508	2.277	E4	11.28
8)	L2 AR-1221-1	1.147	0.963	1.065	1.032	0.993	1.040	E4	6.83
9)	L2 AR-1221-2	8.364	7.018	7.790	7.730	8.017	7.784	E3	6.36
10)	L2 AR-1221-3	2.689	2.264	2.487	2.593	2.758	2.558	E4	7.57
11)	L3 AR-1232-1	1.883					1.883	E4	0.00
12)	L3 AR-1232-2	2.266					2.266	E4	0.00
13)	L3 AR-1232-3	9.545					9.545	E3	0.00
14)	L3 AR-1232-4	7.980					7.980	E3	0.00
15)	L3 AR-1232-5	3.445					3.445	E3	0.00
16)	L4 AR-1242-1	2.089					2.089	E4	0.00
17)	L4 AR-1242-2	4.393					4.393	E4	0.00
18)	L4 AR-1242-3	1.921					1.921	E4	0.00
19)	L4 AR-1242-4	1.743					1.743	E4	0.00
20)	L4 AR-1242-5	2.106					2.106	E4	0.00
21)	L5 AR-1248-1	1.945					1.945	E4	0.00
22)	L5 AR-1248-2	2.836					2.836	E4	0.00
23)	L5 AR-1248-3	2.843					2.843	E4	0.00
24)	L5 AR-1248-4	3.429					3.429	E4	0.00
25)	L5 AR-1248-5	3.407					3.407	E4	0.00
26)	L6 AR-1254-1	4.836					4.836	E4	0.00
27)	L6 AR-1254-2	4.086					4.086	E4	0.00
28)	L6 AR-1254-3	6.621					6.621	E4	0.00
29)	L6 AR-1254-4	4.314					4.314	E4	0.00
30)	L6 AR-1254-5	6.051					6.051	E4	0.00
31)	L7 AR-1260-1	4.523	3.620	4.064	4.565	4.808	4.316	E4	10.96
32)	L7 AR-1260-2	5.942	4.797	5.354	5.975	6.364	5.686	E4	10.80
33)	L7 AR-1260-3	5.317	4.291	4.787	5.303	5.465	5.033	E4	9.69
34)	L7 AR-1260-4	3.680	2.947	3.304	3.682	3.549	3.433	E4	9.09
35)	L7 AR-1260-5	1.030	0.865	0.961	1.011	1.010	0.976	E5	6.83
36)	L8 AR-1262-1	6.707					6.707	E4	0.00
37)	L8 AR-1262-2	1.231					1.231	E5	0.00
38)	L8 AR-1262-3	5.032					5.032	E4	0.00
39)	L8 AR-1262-4	8.845					8.845	E4	0.00
40)	L8 AR-1262-5	3.813					3.813	E4	0.00
41)	L9 AR-1268-1	1.770					1.770	E5	0.00
42)	L9 AR-1268-2	1.593					1.593	E5	0.00
43)	L9 AR-1268-3	1.413					1.413	E5	0.00
44)	L9 AR-1268-4	5.280					5.280	E4	0.00
45)	L9 AR-1268-5	3.975					3.975	E5	0.00

(#) = Out of Range