

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
 Data File : PO099683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 19:40
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 03:42:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 14 03:40:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.719	46797348	34024905	26.305	26.187
2) SA Decachlor...	10.374	8.826	42665605	47188441	25.344	25.599
Target Compounds						
41) L9 AR-1268-1	8.804	7.683	34955796	37327531	257.425	254.981
42) L9 AR-1268-2	8.902	7.747	30137030	33865421	257.749	252.603
43) L9 AR-1268-3	9.141	7.959	28587259	31805408	255.588	253.857
44) L9 AR-1268-4	9.583	8.250	8958177	11126230	247.309	254.365
45) L9 AR-1268-5	10.017	8.556	83077574	89177036	252.603	249.169



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
Data File : PO099683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 19:40
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 03:42:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 14 03:40:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.001 min
Response: 46797348
Conc: 26.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: 0.000 min
Response: 34024905
Conc: 26.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.374 min
Delta R.T.: 0.000 min
Response: 42665605
Conc: 25.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 47188441
Conc: 25.60 ng/ml



#41 AR-1268-1


R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 34955796
Conc: 257.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 37327531
Conc: 254.98 ng/ml




#42 AR-1268-2

R.T.: 8.902 min
Delta R.T.: 0.000 min
Response: 30137030
Conc: 257.75 ng/ml



#42 AR-1268-2


R.T.: 7.747 min
Delta R.T.: 0.000 min
Response: 33865421
Conc: 252.60 ng/ml



#43 AR-1268-3


R.T.: 9.141 min
Delta R.T.: 0.000 min
Response: 28587259
Conc: 255.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.959 min
Delta R.T.: 0.000 min
Response: 31805408
Conc: 253.86 ng/ml




#44 AR-1268-4

R.T.: 9.583 min
Delta R.T.: 0.000 min
Response: 8958177
Conc: 247.31 ng/ml



#44 AR-1268-4


R.T.: 8.250 min
Delta R.T.: -0.002 min
Response: 11126230
Conc: 254.37 ng/ml



#45 AR-1268-5

R.T.: 10.017 min
Delta R.T.: -0.001 min
Response: 83077574
Conc: 252.60 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.556 min
Delta R.T.: 0.000 min
Response: 89177036
Conc: 249.17 ng/ml