

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
 Data File : PO099665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 14:37
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 15:39:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 13 15:27:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.720	131.2E6	92549216	74.759	74.604
2) SA Decachlor...	10.375	8.829	71620422	75846859	74.980	74.855
Target Compounds						
16) L4 AR-1242-1	5.677	4.827	26961031	16363755	746.811	745.457
17) L4 AR-1242-2	5.700	4.847	41420835	22441535	747.346	740.785
18) L4 AR-1242-3	5.763	5.026	27776500	12375372	747.098	747.293
19) L4 AR-1242-4	5.861	5.110	21218403	12792996	745.818	750.283
20) L4 AR-1242-5	6.609	5.641	18652486	14858868	756.496	748.499



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
Data File : PO099665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 14:37
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 15:39:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 13 15:27:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: 0.000 min
Response: 131218742
Conc: 74.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.720 min
Delta R.T.: 0.000 min
Response: 92549216
Conc: 74.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.375 min
Delta R.T.: 0.000 min
Response: 71620422
Conc: 74.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.829 min
Delta R.T.: 0.000 min
Response: 75846859
Conc: 74.85 ng/ml



#16 AR-1242-1


R.T.: 5.677 min
Delta R.T.: 0.000 min
Response: 26961031
Conc: 746.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.827 min
Delta R.T.: 0.000 min
Response: 16363755
Conc: 745.46 ng/ml




#17 AR-1242-2

R.T.: 5.700 min
Delta R.T.: 0.000 min
Response: 41420835
Conc: 747.35 ng/ml



#17 AR-1242-2


R.T.: 4.847 min
Delta R.T.: 0.000 min
Response: 22441535
Conc: 740.79 ng/ml



#18 AR-1242-3


R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 27776500
Conc: 747.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.026 min
Delta R.T.: 0.000 min
Response: 12375372
Conc: 747.29 ng/ml




#19 AR-1242-4

R.T.: 5.861 min
Delta R.T.: 0.000 min
Response: 21218403
Conc: 745.82 ng/ml



#19 AR-1242-4


R.T.: 5.110 min
Delta R.T.: 0.000 min
Response: 12792996
Conc: 750.28 ng/ml



#20 AR-1242-5

R.T.: 6.609 min
Delta R.T.: 0.000 min
Response: 18652486
Conc: 756.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.641 min
Delta R.T.: 0.000 min
Response: 14858868
Conc: 748.50 ng/ml