

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
 Data File : PO099669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 15:44
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 16:40:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 13 16:37:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.721	169.4E6	121.6E6	97.132	97.461
2) SA Decachlor...	10.374	8.828	92614877	99207244	97.858	97.472
Target Compounds						
21) L5 AR-1248-1	5.677	4.826	25641331	15936293	964.468	945.039
22) L5 AR-1248-2	5.955	5.067	39990323	21955792	952.571	948.302
23) L5 AR-1248-3	6.161	5.109	43669801	22843510	959.431	949.365
24) L5 AR-1248-4	6.570	5.285	40324083	26578826	968.953	956.585
25) L5 AR-1248-5	6.610	5.683	39768675	24240622	968.990	964.407


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
Data File : PO099669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 15:44
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 16:40:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 13 16:37:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 169445282
Conc: 97.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.721 min
Delta R.T.: 0.000 min
Response: 121644811
Conc: 97.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.374 min
Delta R.T.: 0.000 min
Response: 92614877
Conc: 97.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.828 min
Delta R.T.: 0.000 min
Response: 99207244
Conc: 97.47 ng/ml



#21 AR-1248-1


R.T.: 5.677 min
Delta R.T.: 0.000 min
Response: 25641331
Conc: 964.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.826 min
Delta R.T.: 0.000 min
Response: 15936293
Conc: 945.04 ng/ml




#22 AR-1248-2

R.T.: 5.955 min
Delta R.T.: 0.000 min
Response: 39990323
Conc: 952.57 ng/ml



#22 AR-1248-2


R.T.: 5.067 min
Delta R.T.: 0.000 min
Response: 21955792
Conc: 948.30 ng/ml



#23 AR-1248-3


R.T.: 6.161 min
Delta R.T.: 0.000 min
Response: 43669801
Conc: 959.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.109 min
Delta R.T.: 0.000 min
Response: 22843510
Conc: 949.37 ng/ml




#24 AR-1248-4

R.T.: 6.570 min
Delta R.T.: 0.000 min
Response: 40324083
Conc: 968.95 ng/ml



#24 AR-1248-4


R.T.: 5.285 min
Delta R.T.: 0.000 min
Response: 26578826
Conc: 956.59 ng/ml



#25 AR-1248-5

R.T.: 6.610 min
Delta R.T.: 0.000 min
Response: 39768675
Conc: 968.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 24240622
Conc: 964.41 ng/ml