

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
 Data File : PO099680.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2023 18:50
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 03:47:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 14 03:47:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.492	3.720	171.0E6	124.7E6	98.027	97.943
2)	SA Decachlor...	10.373	8.828	164.4E6	180.3E6	98.828	98.887
Target Compounds							
41)	L9 AR-1268-1	8.805	7.683	131.6E6	145.1E6	984.481	995.653
42)	L9 AR-1268-2	8.902	7.748	114.7E6	132.5E6	990.262	994.289
43)	L9 AR-1268-3	9.140	7.959	108.8E6	123.7E6	986.404	993.671
44)	L9 AR-1268-4	9.582	8.251	35644617	44631177	991.957	1010.071
45)	L9 AR-1268-5	10.017	8.556	328.5E6	362.5E6	999.398	1006.359



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111323\
Data File : PO099680.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2023 18:50
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 03:47:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 14 03:47:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 171019018
Conc: 98.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.720 min
Delta R.T.: 0.000 min
Response: 124690699
Conc: 97.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.373 min
Delta R.T.: 0.000 min
Response: 164446087
Conc: 98.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.828 min
Delta R.T.: 0.000 min
Response: 180278284
Conc: 98.89 ng/ml



#41 AR-1268-1


R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 131639641
Conc: 984.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 145126348
Conc: 995.65 ng/ml




#42 AR-1268-2

R.T.: 8.902 min
Delta R.T.: 0.000 min
Response: 114668855
Conc: 990.26 ng/ml



#42 AR-1268-2


R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 132542852
Conc: 994.29 ng/ml



#43 AR-1268-3


R.T.: 9.140 min
Delta R.T.: 0.000 min
Response: 108848119
Conc: 986.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.959 min
Delta R.T.: 0.000 min
Response: 123712864
Conc: 993.67 ng/ml




#44 AR-1268-4

R.T.: 9.582 min
Delta R.T.: 0.000 min
Response: 35644617
Conc: 991.96 ng/ml



#44 AR-1268-4


R.T.: 8.251 min
Delta R.T.: 0.000 min
Response: 44631177
Conc: 1010.07 ng/ml



#45 AR-1268-5

R.T.: 10.017 min
Delta R.T.: 0.000 min
Response: 328489855
Conc: 999.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.556 min
Delta R.T.: 0.000 min
Response: 362478568
Conc: 1006.36 ng/ml