

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111518\
 Data File : P0051566.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Nov 2018 16:30
 Operator : SM/SJ
 Sample : J5767-13
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CL-01-111418-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 00:04:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 15 08:04:19 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.308	3.298	43052958	16446795	13.481	14.936
2) SA Decachlor...	9.949	8.053	23272258	9022253	12.125	10.884

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111518\
Data File : PO051566.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Nov 2018 16:30
Operator : SM/SJ
Sample : J5767-13
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-01-111418-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 00:04:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111318.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 15 08:04:19 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.308 min
Delta R.T.: -0.005 min
Response: 43052958
Conc: 13.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
CL-01-111418-C




#1 Tetrachloro-m-xylene

R.T.: 3.298 min
Delta R.T.: -0.004 min
Response: 16446795
Conc: 14.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.949 min
Delta R.T.: -0.012 min
Response: 23272258
Conc: 12.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.053 min
Delta R.T.: -0.006 min
Response: 9022253
Conc: 10.88 ng/ml