

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111519\
 Data File : P0063727.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Nov 2019 10:31
 Operator : SM/AJ
 Sample : PB124825BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB124825BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 11:54:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.569	733137	518475	23.978	23.343
2)	SA Decachlor...	9.939	8.540	935502	676474	23.022	23.130
Target Compounds							
3)	L1 AR-1016-1	5.476	4.648	350232	247983	252.182	241.988
4)	L1 AR-1016-2	5.498	4.667	497673	335536	253.561	246.273
5)	L1 AR-1016-3	5.559	4.841	320871	181993	256.868	244.409
6)	L1 AR-1016-4	5.658	4.882	262414	155283	255.139	249.247
7)	L1 AR-1016-5	5.950	5.094	280377	199637	263.491	251.361
31)	L7 AR-1260-1	7.069	6.122	546461	402015	270.779	255.994
32)	L7 AR-1260-2	7.326	6.309	658018	502996	261.182	255.701
33)	L7 AR-1260-3	7.683	6.463	520818	455771	257.047	255.550
34)	L7 AR-1260-4	7.907	6.933	608002	392305	255.305	249.168
35)	L7 AR-1260-5	8.218	7.173	1109056	908049	236.256	234.218



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111519\
Data File : PO063727.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Nov 2019 10:31
Operator : SM/AJ
Sample : PB124825BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB124825BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 11:54:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111419.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 15 01:29:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.002 min
Response: 733137
Conc: 23.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#1 Tetrachloro-m-xylene

R.T.: 3.569 min
Delta R.T.: 0.000 min
Response: 518475
Conc: 23.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.939 min
Delta R.T.: 0.004 min
Response: 935502
Conc: 23.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.540 min
Delta R.T.: 0.000 min
Response: 676474
Conc: 23.13 ng/ml



#3 AR-1016-1


R.T.: 5.476 min
Delta R.T.: 0.000 min
Response: 350232
Conc: 252.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#3 AR-1016-1

R.T.: 4.648 min
Delta R.T.: 0.000 min
Response: 247983
Conc: 241.99 ng/ml




#4 AR-1016-2

R.T.: 5.498 min
Delta R.T.: 0.001 min
Response: 497673
Conc: 253.56 ng/ml



#4 AR-1016-2


R.T.: 4.667 min
Delta R.T.: 0.000 min
Response: 335536
Conc: 246.27 ng/ml



#5 AR-1016-3


R.T.: 5.559 min
Delta R.T.: 0.001 min
Response: 320871
Conc: 256.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#5 AR-1016-3

R.T.: 4.841 min
Delta R.T.: 0.000 min
Response: 181993
Conc: 244.41 ng/ml




#6 AR-1016-4

R.T.: 5.658 min
Delta R.T.: 0.002 min
Response: 262414
Conc: 255.14 ng/ml



#6 AR-1016-4


R.T.: 4.882 min
Delta R.T.: 0.000 min
Response: 155283
Conc: 249.25 ng/ml



#7 AR-1016-5


R.T.: 5.950 min
Delta R.T.: 0.000 min
Response: 280377
Conc: 263.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#7 AR-1016-5

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 199637
Conc: 251.36 ng/ml




#31 AR-1260-1

R.T.: 7.069 min
Delta R.T.: 0.001 min
Response: 546461
Conc: 270.78 ng/ml



#31 AR-1260-1


R.T.: 6.122 min
Delta R.T.: 0.000 min
Response: 402015
Conc: 255.99 ng/ml



#32 AR-1260-2


R.T.: 7.326 min
Delta R.T.: 0.000 min
Response: 658018
Conc: 261.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#32 AR-1260-2

R.T.: 6.309 min
Delta R.T.: 0.001 min
Response: 502996
Conc: 255.70 ng/ml




#33 AR-1260-3

R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 520818
Conc: 257.05 ng/ml



#33 AR-1260-3


R.T.: 6.463 min
Delta R.T.: 0.000 min
Response: 455771
Conc: 255.55 ng/ml



#34 AR-1260-4


R.T.: 7.907 min
Delta R.T.: 0.002 min
Response: 608002
Conc: 255.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124825BS




#34 AR-1260-4

R.T.: 6.933 min
Delta R.T.: 0.000 min
Response: 392305
Conc: 249.17 ng/ml



#35 AR-1260-5

R.T.: 8.218 min
Delta R.T.: 0.002 min
Response: 1109056
Conc: 236.26 ng/ml



#35 AR-1260-5

R.T.: 7.173 min
Delta R.T.: 0.000 min
Response: 908049
Conc: 234.22 ng/ml