

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111519\
 Data File : P0063772.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Nov 2019 23:47
 Operator : SM/AJ
 Sample : K5838-01
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 UC-TP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 09:34:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.569	1385930	1033085	45.329	46.512
2)	SA Decachlor...	9.936	8.538	1562923	1045461	38.463	35.746
Target Compounds							
21)	L5 AR-1248-1	5.477	4.647	613211	445312	702.155	691.052
22)	L5 AR-1248-2	5.748	4.881	2771008	1903831	2186.977	2208.914
23)	L5 AR-1248-3	5.950	4.923	3097766	1427478	2171.509	1613.428 #
24)	L5 AR-1248-4	6.353	5.094	4258540	1648749	2460.313	1531.525 #
25)	L5 AR-1248-5	6.390	5.484	4952010	2293088	2936.009	2055.930 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111519\
Data File : PO063772.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Nov 2019 23:47
Operator : SM/AJ
Sample : K5838-01
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
UC-TP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 09:34:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111419.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 15 01:29:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: 0.003 min
Response: 1385930
Conc: 45.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
UC-TP




#1 Tetrachloro-m-xylene

R.T.: 3.569 min
Delta R.T.: 0.000 min
Response: 1033085
Conc: 46.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: 0.000 min
Response: 1562923
Conc: 38.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.538 min
Delta R.T.: -0.001 min
Response: 1045461
Conc: 35.75 ng/ml



#21 AR-1248-1


R.T.: 5.477 min
Delta R.T.: 0.001 min
Response: 613211
Conc: 702.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
UC-TP




#21 AR-1248-1

R.T.: 4.647 min
Delta R.T.: 0.000 min
Response: 445312
Conc: 691.05 ng/ml




#22 AR-1248-2

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 2771008
Conc: 2186.98 ng/ml



#22 AR-1248-2


R.T.: 4.881 min
Delta R.T.: -0.001 min
Response: 1903831
Conc: 2208.91 ng/ml



#23 AR-1248-3


R.T.: 5.950 min
Delta R.T.: 0.000 min
Response: 3097766
Conc: 2171.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
UC-TP




#23 AR-1248-3

R.T.: 4.923 min
Delta R.T.: -0.001 min
Response: 1427478
Conc: 1613.43 ng/ml




#24 AR-1248-4

R.T.: 6.353 min
Delta R.T.: 0.001 min
Response: 4258540
Conc: 2460.31 ng/ml



#24 AR-1248-4


R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 1648749
Conc: 1531.53 ng/ml



#25 AR-1248-5

R.T.: 6.390 min
Delta R.T.: 0.000 min
Response: 4952010
Conc: 2936.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
UC-TP



#25 AR-1248-5

R.T.: 5.484 min
Delta R.T.: -0.001 min
Response: 2293088
Conc: 2055.93 ng/ml