

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111523\
 Data File : PO099771.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2023 09:58
 Operator : YP/AJ
 Sample : 05382-01 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 S-11

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/16/2023
 Supervised By :Ankita Jodhani 11/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 11:35:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 14 04:21:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.487	3.717	4187718	3093613	2.449	2.619
2)	SA Decachlor...	10.367	8.817	2140256	2361629	2.292	2.406
Target Compounds							
16)	L4 AR-1242-1	5.673	4.821	20345490	12112771	555.360	527.266m
17)	L4 AR-1242-2	5.696	4.841	29419979	15900766	519.408	506.166m
18)	L4 AR-1242-3	5.759	5.020	21247846	9422946	560.717	563.046m
19)	L4 AR-1242-4	5.858	5.104	16782201	8821677	585.161	510.530m
20)	L4 AR-1242-5	6.604	5.634	18252391	14343506	742.632m	720.972m
31)	L7 AR-1260-1	7.296	6.325	2649614	4065956	47.486m	84.462m#
32)	L7 AR-1260-2	7.556	6.513	2556401	3539842	42.376	62.418 #
33)	L7 AR-1260-3	7.920	6.671	1438431	2567357	34.414	46.444 #
34)	L7 AR-1260-4	8.148	7.147	1781423	1514241	36.174m	34.360
35)	L7 AR-1260-5	8.479	7.389	2197661	2998611	26.192	30.172

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111523\
Data File : PO099771.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2023 09:58
Operator : YP/AJ
Sample : 05382-01 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-11

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/16/2023
Supervised By :Ankita Jodhani 11/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 11:35:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 14 04:21:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

