

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111622\
 Data File : P0090634.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Nov 2022 09:33
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 20:26:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	

System Monitoring Compounds								
1)	SA Tetrachlo...	4.341f	0.000	404546	0	0.126	N.D.	#
Target Compounds								
3)	L1 AR-1016-1	5.414f	0.000	516125	0	4.852	N.D.	#
16)	L4 AR-1242-1	5.414f	0.000	516125	0	6.238	N.D.	#
20)	L4 AR-1242-5	6.345f	0.000	73414	0	1.194	N.D.	#
21)	L5 AR-1248-1	5.414f	0.000	516125	0	8.160	N.D.	#
24)	L5 AR-1248-4	6.345	0.000	73414	0	0.689	N.D.	#
25)	L5 AR-1248-5	6.345f	0.000	73414	0	0.699	N.D.	#
26)	L6 AR-1254-1	6.345	0.000	73414	0	0.633	N.D.	#
28)	L6 AR-1254-3	6.926	0.000	290745	0	1.675	N.D.	#
30)	L6 AR-1254-5	7.692f	0.000	139935	0	0.993	N.D.	#
32)	L7 AR-1260-2	7.311f	0.000	187249	0	1.157	N.D.	#
33)	L7 AR-1260-3	7.715	0.000	80317	0	0.765	N.D.	#
34)	L7 AR-1260-4	7.932	0.000	113731	0	0.921	N.D.	#
36)	L8 AR-1262-1	7.715	0.000	80317	0	0.485	N.D.	#
38)	L8 AR-1262-3	8.625f	0.000	229830	0	1.170	N.D.	#
39)	L8 AR-1262-4	8.625	7.531f	229830	51550	2.362	0.531	#
41)	L9 AR-1268-1	8.625f	0.000	229830	0	0.659	N.D.	#
42)	L9 AR-1268-2	8.625	7.531f	229830	51550	0.737	0.367	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111622\
Data File : PO090634.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Nov 2022 09:33
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 20:26:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.341 min
Delta R.T.: 0.034 min
Response: 404546
Conc: 0.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.487 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.054 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.496 min
Response: 0
Conc: N.D.



#3 AR-1016-1


R.T.: 5.414 min
Delta R.T.: -0.068 min
Response: 516125
Conc: 4.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.574 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.414 min
Delta R.T.: -0.067 min
Response: 516125
Conc: 6.24 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.572 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.345 min
Delta R.T.: -0.061 min
Response: 73414
Conc: 1.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.377 min
Response: 0
Conc: N.D.




#21 AR-1248-1

R.T.: 5.414 min
Delta R.T.: -0.066 min
Response: 516125
Conc: 8.16 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T. : 4.573 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.345 min
Delta R.T.: -0.022 min
Response: 73414
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.023 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.345 min
Delta R.T.: -0.060 min
Response: 73414
Conc: 0.70 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 5.418 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.345 min
Delta R.T.: 0.005 min
Response: 73414
Conc: 0.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.376 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.926 min
Delta R.T.: -0.002 min
Response: 290745
Conc: 1.67 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.930 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.692 min
Delta R.T.: 0.055 min
Response: 139935
Conc: 0.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.579 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.311 min
Delta R.T.: -0.043 min
Response: 187249
Conc: 1.16 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.253 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.715 min
Delta R.T.: -0.001 min
Response: 80317
Conc: 0.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.406 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.932 min
Delta R.T.: -0.010 min
Response: 113731
Conc: 0.92 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.881 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.715 min
Delta R.T.: 0.000 min
Response: 80317
Conc: 0.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 6.618 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.625 min
Delta R.T.: 0.058 min
Response: 229830
Conc: 1.17 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.407 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.625 min
Delta R.T.: -0.029 min
Response: 229830
Conc: 2.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#39 AR-1262-4

R.T.: 7.531 min
Delta R.T.: 0.061 min
Response: 51550
Conc: 0.53 ng/ml




#41 AR-1268-1

R.T.: 8.625 min
Delta R.T.: 0.062 min
Response: 229830
Conc: 0.66 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 7.406 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.625 min
Delta R.T.: -0.030 min
Response: 229830
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE



#42 AR-1268-2

R.T.: 7.531 min
Delta R.T.: 0.060 min
Response: 51550
Conc: 0.37 ng/ml