

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111720\
 Data File : PO073270.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Nov 2020 18:35
 Operator : DD\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Ankita
 11/18/2020 10:24:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 18 03:46:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 18 03:34:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.945	4.015	330495	195999	5.088	5.316
2) SA Decachlor...	10.976	9.289	331345	219076	5.550	5.718
Target Compounds						
21) L5 AR-1248-1	6.268	5.267	78426	52225	52.218m	55.462
22) L5 AR-1248-2	6.564	5.529	104784	73028	52.590m	57.667
23) L5 AR-1248-3	6.782	5.574	109285	76251	48.763	57.977
24) L5 AR-1248-4	7.212	5.757	165828	92665	55.783	58.536
25) L5 AR-1248-5	7.252	6.177	149925	93513	53.681	56.392

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111720\
Data File : PO073270.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Nov 2020 18:35
Operator : DD\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
11/18/2020 10:24:07 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 03:46:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111720.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 18 03:34:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.945 min
Delta R.T.: 0.000 min
Response: 330495
Conc: 5.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
11/18/2020 10:24:07 AM




#1 Tetrachloro-m-xylene

R.T.: 4.015 min
Delta R.T.: 0.000 min
Response: 195999
Conc: 5.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.976 min
Delta R.T.: 0.000 min
Response: 331345
Conc: 5.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.289 min
Delta R.T.: 0.000 min
Response: 219076
Conc: 5.72 ng/ml




#21 AR-1248-1

R.T.: 6.268 min
Delta R.T.: 0.000 min
Response: 78426
Conc: 52.22 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
11/18/2020 10:24:07 AM




#21 AR-1248-1

R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 52225
Conc: 55.46 ng/ml




#22 AR-1248-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 104784
Conc: 52.59 ng/ml m



#22 AR-1248-2

R.T.: 5.529 min
Delta R.T.: 0.000 min
Response: 73028
Conc: 57.67 ng/ml




#23 AR-1248-3

R.T.: 6.782 min
Delta R.T.: 0.000 min
Response: 109285
Conc: 48.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
11/18/2020 10:24:07 AM




#23 AR-1248-3

R.T.: 5.574 min
Delta R.T.: 0.000 min
Response: 76251
Conc: 57.98 ng/ml




#24 AR-1248-4

R.T.: 7.212 min
Delta R.T.: 0.000 min
Response: 165828
Conc: 55.78 ng/ml



#24 AR-1248-4

R.T.: 5.757 min
Delta R.T.: 0.000 min
Response: 92665
Conc: 58.54 ng/ml




#25 AR-1248-5

R.T.: 7.252 min
Delta R.T.: 0.000 min
Response: 149925
Conc: 53.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Ankita
11/18/2020 10:24:07 AM



#25 AR-1248-5

R.T.: 6.177 min
Delta R.T.: 0.000 min
Response: 93513
Conc: 56.39 ng/ml