

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111723\
 Data File : PO099934.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2023 18:41
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 17 21:46:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 14 04:21:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.490	3.715	82663032	65357928	48.338	55.331
2)	SA Decachlor...	10.373	8.813	50372581	55035519	53.951	56.070
Target Compounds							
16)	L4 AR-1242-1	5.675	4.820	18433006	11849735	503.156	515.817
17)	L4 AR-1242-2	5.698	4.839	27413771	16105232	483.989	512.675
18)	L4 AR-1242-3	5.761	5.019	18122285	8940353	478.235	534.210
19)	L4 AR-1242-4	5.861	5.103	13968535	9260090	487.054	535.902
20)	L4 AR-1242-5	6.607	5.633	13071285	11359932	531.829	571.003

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111723\
Data File : PO099934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2023 18:41
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 17 21:46:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 14 04:21:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

