

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111723\
 Data File : P0099929.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Nov 2023 16:49
 Operator : YP/AJ
 Sample : 05467-04 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KZZ372H-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/20/2023
 Supervised By :Ankita Jodhani 11/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 17 21:44:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 14 04:21:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.489	3.715	2738443	2178280	1.601	1.844
2) SA Decachlor...	10.372	8.814	1566407	2020432	1.678m	2.058

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111723\
Data File : PO099929.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2023 16:49
Operator : YP/AJ
Sample : 05467-04 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
KZZ372H-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/20/2023
Supervised By :Ankita Jodhani 11/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 17 21:44:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 14 04:21:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.489 min
Delta R.T.: -0.002 min
Response: 2738443
Conc: 1.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
KZZ372H-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/20/2023
Supervised By :Ankita Jodhani 11/20/2023




#1 Tetrachloro-m-xylene

R.T.: 3.715 min
Delta R.T.: -0.005 min
Response: 2178280
Conc: 1.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.372 min
Delta R.T.: 0.000 min
Response: 1566407
Conc: 1.68 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.814 min
Delta R.T.: -0.015 min
Response: 2020432
Conc: 2.06 ng/ml