

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111819\
 Data File : P0063828.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Nov 2019 12:13
 Operator : SM/AJ
 Sample : K5903-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EAST-END-EXTRA-SAMPLE

Manual Integrations
 APPROVED

Ankita
 11/19/2019 3:59:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 01:12:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.567	670843	537926	21.941	24.219
2)	SA Decachlor...	9.944	8.540	791939	620453	19.489	21.214m
Target Compounds							
31)	L7 AR-1260-1	7.072	6.121	31925	40779	15.819	25.967m#
32)	L7 AR-1260-2	7.327	6.308	49267	48953	19.555	24.886m#
33)	L7 AR-1260-3	7.687	6.461	36339	45557	17.935	25.544m#
34)	L7 AR-1260-4	7.908	6.931	59539	36400	25.001	23.119m
35)	L7 AR-1260-5	8.221	7.172	119519	114943	25.460	29.648m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111819\
Data File : PO063828.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Nov 2019 12:13
Operator : SM/AJ
Sample : K5903-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EAST-END-EXTRA-SAMPLE

Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 01:12:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111419.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 15 01:29:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

