

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111819\
 Data File : P0063861.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Nov 2019 22:08
 Operator : SM/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 11/19/2019 3:59:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 01:20:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.568	1437914	1052072	47.029	47.366
2)	SA Decachlor...	9.946	8.540	1801095	1503871	44.324	51.420
Target Compounds							
3)	L1 AR-1016-1	5.479	4.646	684918	526545	493.169	513.814
4)	L1 AR-1016-2	5.500	4.666	976382	724399	497.459	531.685
5)	L1 AR-1016-3	5.562	4.840	620728	395061	496.912	530.551
6)	L1 AR-1016-4	5.661	4.880	510968	329669	496.802	529.158
7)	L1 AR-1016-5	5.952	5.092	545127	432700	512.298	544.807m
31)	L7 AR-1260-1	7.073	6.121	1051215	866638	520.892	551.855
32)	L7 AR-1260-2	7.330	6.308	1277789	1083506	507.183	550.808
33)	L7 AR-1260-3	7.687	6.461	1002354	1002492	494.706	562.097
34)	L7 AR-1260-4	7.911	6.932	1155583	877624	485.238	557.412
35)	L7 AR-1260-5	8.223	7.172	2188724	2125403	466.252	548.216

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111819\
Data File : PO063861.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Nov 2019 22:08
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 01:20:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111419.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 15 01:29:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.317 min
Delta R.T.: 0.003 min
Response: 1437914
Conc: 47.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#1 Tetrachloro-m-xylene

R.T.: 3.568 min
Delta R.T.: 0.000 min
Response: 1052072
Conc: 47.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.946 min
Delta R.T.: 0.010 min
Response: 1801095
Conc: 44.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.540 min
Delta R.T.: 0.000 min
Response: 1503871
Conc: 51.42 ng/ml




#3 AR-1016-1

R.T.: 5.479 min
Delta R.T.: 0.003 min
Response: 684918
Conc: 493.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#3 AR-1016-1

R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 526545
Conc: 513.81 ng/ml




#4 AR-1016-2

R.T.: 5.500 min
Delta R.T.: 0.003 min
Response: 976382
Conc: 497.46 ng/ml



#4 AR-1016-2

R.T.: 4.666 min
Delta R.T.: 0.000 min
Response: 724399
Conc: 531.68 ng/ml




#5 AR-1016-3

R.T.: 5.562 min
Delta R.T.: 0.003 min
Response: 620728
Conc: 496.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#5 AR-1016-3

R.T.: 4.840 min
Delta R.T.: -0.001 min
Response: 395061
Conc: 530.55 ng/ml




#6 AR-1016-4

R.T.: 5.661 min
Delta R.T.: 0.004 min
Response: 510968
Conc: 496.80 ng/ml



#6 AR-1016-4

R.T.: 4.880 min
Delta R.T.: -0.001 min
Response: 329669
Conc: 529.16 ng/ml




#7 AR-1016-5

R.T.: 5.952 min
Delta R.T.: 0.003 min
Response: 545127
Conc: 512.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#7 AR-1016-5

R.T.: 5.092 min
Delta R.T.: -0.002 min
Response: 432700
Conc: 544.81 ng/ml m




#31 AR-1260-1

R.T.: 7.073 min
Delta R.T.: 0.005 min
Response: 1051215
Conc: 520.89 ng/ml



#31 AR-1260-1

R.T.: 6.121 min
Delta R.T.: 0.000 min
Response: 866638
Conc: 551.86 ng/ml




#32 AR-1260-2

R.T.: 7.330 min
Delta R.T.: 0.005 min
Response: 1277789
Conc: 507.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#32 AR-1260-2

R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 1083506
Conc: 550.81 ng/ml




#33 AR-1260-3

R.T.: 7.687 min
Delta R.T.: 0.005 min
Response: 1002354
Conc: 494.71 ng/ml



#33 AR-1260-3

R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 1002492
Conc: 562.10 ng/ml




#34 AR-1260-4

R.T.: 7.911 min
Delta R.T.: 0.006 min
Response: 1155583
Conc: 485.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
11/19/2019 3:59:48 PM




#34 AR-1260-4

R.T.: 6.932 min
Delta R.T.: 0.000 min
Response: 877624
Conc: 557.41 ng/ml



#35 AR-1260-5

R.T.: 8.223 min
Delta R.T.: 0.007 min
Response: 2188724
Conc: 466.25 ng/ml



#35 AR-1260-5

R.T.: 7.172 min
Delta R.T.: 0.000 min
Response: 2125403
Conc: 548.22 ng/ml