

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0111919\
 Data File : P0063869.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Nov 2019 10:00
 Operator : SM/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 10:44:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.567	1299442	959223	42.500	43.186
2)	SA Decachlor...	9.945	8.538	1778847	1526613	43.777	52.198
Target Compounds							
3)	L1 AR-1016-1	5.478	4.645	649403	497201	467.597	485.180
4)	L1 AR-1016-2	5.500	4.664	931023	677613	474.349	497.346
5)	L1 AR-1016-3	5.562	4.838	598284	373764	478.945	501.950
6)	L1 AR-1016-4	5.660	4.879	494881	313838	481.161	503.747
7)	L1 AR-1016-5	5.952	5.090	522013	415920	490.575	523.680m
31)	L7 AR-1260-1	7.071	6.119	1068860	890823	529.635	567.256
32)	L7 AR-1260-2	7.328	6.306	1297383	1120373	514.960	569.549
33)	L7 AR-1260-3	7.686	6.459	1013838	1018440	500.373	571.039
34)	L7 AR-1260-4	7.910	6.930	1178732	880151	494.959	559.018
35)	L7 AR-1260-5	8.221	7.170	2209121	2109788	470.597	544.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111919\
Data File : PO063869.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Nov 2019 10:00
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 10:44:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111419.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 15 01:29:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.002 min
Response: 1299442
Conc: 42.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: -0.002 min
Response: 959223
Conc: 43.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.945 min
Delta R.T.: 0.010 min
Response: 1778847
Conc: 43.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.538 min
Delta R.T.: -0.001 min
Response: 1526613
Conc: 52.20 ng/ml



#3 AR-1016-1


R.T.: 5.478 min
Delta R.T.: 0.003 min
Response: 649403
Conc: 467.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.645 min
Delta R.T.: -0.003 min
Response: 497201
Conc: 485.18 ng/ml




#4 AR-1016-2

R.T.: 5.500 min
Delta R.T.: 0.003 min
Response: 931023
Conc: 474.35 ng/ml



#4 AR-1016-2


R.T.: 4.664 min
Delta R.T.: -0.002 min
Response: 677613
Conc: 497.35 ng/ml



#5 AR-1016-3


R.T.: 5.562 min
Delta R.T.: 0.003 min
Response: 598284
Conc: 478.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.838 min
Delta R.T.: -0.003 min
Response: 373764
Conc: 501.95 ng/ml




#6 AR-1016-4

R.T.: 5.660 min
Delta R.T.: 0.004 min
Response: 494881
Conc: 481.16 ng/ml



#6 AR-1016-4


R.T.: 4.879 min
Delta R.T.: -0.002 min
Response: 313838
Conc: 503.75 ng/ml



#7 AR-1016-5


R.T.: 5.952 min
Delta R.T.: 0.003 min
Response: 522013
Conc: 490.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.090 min
Delta R.T.: -0.003 min
Response: 415920
Conc: 523.68 ng/ml m




#31 AR-1260-1

R.T.: 7.071 min
Delta R.T.: 0.004 min
Response: 1068860
Conc: 529.63 ng/ml



#31 AR-1260-1


R.T.: 6.119 min
Delta R.T.: -0.003 min
Response: 890823
Conc: 567.26 ng/ml



#32 AR-1260-2


R.T.: 7.328 min
Delta R.T.: 0.003 min
Response: 1297383
Conc: 514.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.306 min
Delta R.T.: -0.002 min
Response: 1120373
Conc: 569.55 ng/ml




#33 AR-1260-3

R.T.: 7.686 min
Delta R.T.: 0.004 min
Response: 1013838
Conc: 500.37 ng/ml



#33 AR-1260-3


R.T.: 6.459 min
Delta R.T.: -0.002 min
Response: 1018440
Conc: 571.04 ng/ml



#34 AR-1260-4


R.T.: 7.910 min
Delta R.T.: 0.005 min
Response: 1178732
Conc: 494.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.930 min
Delta R.T.: -0.003 min
Response: 880151
Conc: 559.02 ng/ml



#35 AR-1260-5

R.T.: 8.221 min
Delta R.T.: 0.006 min
Response: 2209121
Conc: 470.60 ng/ml



#35 AR-1260-5

R.T.: 7.170 min
Delta R.T.: -0.002 min
Response: 2109788
Conc: 544.19 ng/ml