

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112025\
 Data File : PO115255.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2025 10:51
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/21/2025
 Supervised By :Sohil Jodhani 12/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 12:50:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.643	386.5E6	240.0E6	54.913	53.730
2) SA Decachlor...	9.917	8.629	245.9E6	222.1E6	55.633	56.905
Target Compounds						
3) L1 AR-1016-1	5.479	4.719	133.0E6	105.8E6	540.562	556.225m
4) L1 AR-1016-2	5.501	4.738	212.9E6	138.3E6	536.054	510.611m
5) L1 AR-1016-3	5.562	4.913	123.3E6	77983466	552.903	527.011
6) L1 AR-1016-4	5.657	4.955	105.6E6	61224348	559.360	521.423
7) L1 AR-1016-5	5.949	5.167	92892875	81422171	521.293	533.169
31) L7 AR-1260-1	7.063	6.196	158.2E6	145.2E6	517.645	490.362
32) L7 AR-1260-2	7.316	6.384	222.8E6	204.7E6	562.738	476.008
33) L7 AR-1260-3	7.674	6.536	185.7E6	171.8E6	559.688	489.220
34) L7 AR-1260-4	7.895	7.007	165.5E6	151.5E6	542.465	500.090
35) L7 AR-1260-5	8.202	7.249	392.0E6	418.1E6	549.577	508.875

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112025\
Data File : PO115255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2025 10:51
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 12:50:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

