

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO112025\
 Data File : PO115274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2025 18:36
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/21/2025
 Supervised By :Sohil Jodhani 12/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 01:53:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.341	3.644	375.0E6	234.9E6	53.278	52.583
2) SA Decachlor...	9.918	8.630	242.4E6	214.4E6	54.840	54.933
Target Compounds						
26) L6 AR-1254-1	6.320	5.518	102.4E6	140.7E6	404.025m	468.555m
27) L6 AR-1254-2	6.537	5.665	187.8E6	119.6E6	506.909m	443.704m
28) L6 AR-1254-3	6.897	6.067	199.3E6	195.5E6	517.872m	463.026
29) L6 AR-1254-4	7.179	6.296	193.0E6	133.3E6	587.129	448.093
30) L6 AR-1254-5	7.594	6.712	137.0E6	178.2E6	453.923m	448.395

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112025\
Data File : PO115274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2025 18:36
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 01:53:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.341 min
Delta R.T.: -0.007 min
Response: 374977215
Conc: 53.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.004 min
Response: 234878009
Conc: 52.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.918 min
Delta R.T.: -0.012 min
Response: 242405086
Conc: 54.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.630 min
Delta R.T.: -0.010 min
Response: 214448808
Conc: 54.93 ng/ml




#26 AR-1254-1

R.T.: 6.320 min
Delta R.T.: -0.011 min
Response: 102391804
Conc: 404.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025




#26 AR-1254-1

R.T.: 5.518 min
Delta R.T.: -0.009 min
Response: 140650867
Conc: 468.55 ng/ml m




#27 AR-1254-2

R.T.: 6.537 min
Delta R.T.: -0.010 min
Response: 187781495
Conc: 506.91 ng/ml m



#27 AR-1254-2

R.T.: 5.665 min
Delta R.T.: -0.009 min
Response: 119625982
Conc: 443.70 ng/ml m




#28 AR-1254-3

R.T.: 6.897 min
Delta R.T.: -0.010 min
Response: 199310441
Conc: 517.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025




#28 AR-1254-3

R.T.: 6.067 min
Delta R.T.: -0.008 min
Response: 195510254
Conc: 463.03 ng/ml




#29 AR-1254-4

R.T.: 7.179 min
Delta R.T.: -0.009 min
Response: 192998793
Conc: 587.13 ng/ml



#29 AR-1254-4

R.T.: 6.296 min
Delta R.T.: -0.008 min
Response: 133260188
Conc: 448.09 ng/ml




#30 AR-1254-5

R.T.: 7.594 min
Delta R.T.: -0.010 min
Response: 137018292
Conc: 453.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/21/2025
Supervised By :Sohil Jodhani 12/01/2025



#30 AR-1254-5

R.T.: 6.712 min
Delta R.T.: -0.008 min
Response: 178248853
Conc: 448.40 ng/ml