

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112122\
 Data File : PO090855.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2022 23:53
 Operator : YP/AJ
 Sample : N5733-08 10X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 04:29:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.298	3.477	4149970	1675996	1.291	1.577
2)	SA Decachlor...	10.046	8.479	4764190	1650616	1.944	1.603
Target Compounds							
21)	L5 AR-1248-1	5.411f	0.000	3196487	0	50.540	N.D. #
28)	L6 AR-1254-3	6.921	0.000	620658	0	3.575	N.D. #
29)	L6 AR-1254-4	7.203	0.000	236139	0	1.853	N.D. #
30)	L6 AR-1254-5	7.706f	0.000	645252	0	4.581	N.D. #
31)	L7 AR-1260-1	7.088	0.000	299774	0	2.138	N.D. #
32)	L7 AR-1260-2	7.351	0.000	127577	0	0.789	N.D. #
33)	L7 AR-1260-3	7.706	0.000	645252	0	6.142	N.D. #
34)	L7 AR-1260-4	7.937	0.000	233691	0	1.892	N.D. #
35)	L7 AR-1260-5	8.258	0.000	1248486	0	5.486	N.D. #
36)	L8 AR-1262-1	7.706	0.000	645252	0	3.892	N.D. #
37)	L8 AR-1262-2	8.258	0.000	1248486	0	4.444	N.D. #
38)	L8 AR-1262-3	8.557	0.000	545412	0	2.778	N.D. #
39)	L8 AR-1262-4	8.649	0.000	251352	0	2.583	N.D. #
40)	L8 AR-1262-5	9.373f	0.000	721024	0	6.843	N.D. #
41)	L9 AR-1268-1	8.557	0.000	545412	0	1.564	N.D. #
42)	L9 AR-1268-2	8.649	0.000	251352	0	0.806	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112122\
Data File : PO090855.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2022 23:53
Operator : YP/AJ
Sample : N5733-08 10X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 04:29:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: -0.009 min
Response: 4149970
Conc: 1.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.477 min
Delta R.T.: -0.010 min
Response: 1675996
Conc: 1.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.008 min
Response: 4764190
Conc: 1.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.479 min
Delta R.T.: -0.017 min
Response: 1650616
Conc: 1.60 ng/ml



#21 AR-1248-1


R.T.: 5.411 min
Delta R.T.: -0.069 min
Response: 3196487
Conc: 50.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.573 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.921 min
Delta R.T.: -0.007 min
Response: 620658
Conc: 3.58 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.930 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.203 min
Delta R.T.: -0.011 min
Response: 236139
Conc: 1.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.160 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.706 min
Delta R.T.: 0.069 min
Response: 645252
Conc: 4.58 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.579 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.088 min
Delta R.T.: -0.007 min
Response: 299774
Conc: 2.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.064 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.351 min
Delta R.T.: -0.003 min
Response: 127577
Conc: 0.79 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.253 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.706 min
Delta R.T.: -0.009 min
Response: 645252
Conc: 6.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.406 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.937 min
Delta R.T.: -0.004 min
Response: 233691
Conc: 1.89 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.881 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.258 min
Delta R.T.: 0.000 min
Response: 1248486
Conc: 5.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.125 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.706 min
Delta R.T.: -0.008 min
Response: 645252
Conc: 3.89 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.618 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.258 min
Delta R.T.: 0.001 min
Response: 1248486
Conc: 4.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.121 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.557 min
Delta R.T.: -0.009 min
Response: 545412
Conc: 2.78 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.407 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.649 min
Delta R.T.: -0.005 min
Response: 251352
Conc: 2.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.470 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.373 min
Delta R.T.: 0.070 min
Response: 721024
Conc: 6.84 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 7.966 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.557 min
Delta R.T.: -0.005 min
Response: 545412
Conc: 1.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.406 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.649 min
Delta R.T.: -0.006 min
Response: 251352
Conc: 0.81 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.471 min
Response: 0
Conc: N.D.